

**FoundationOne® Liquid CDx  
Technical Information**
**Table of Contents**

<b>1</b>	<b>Intended Use</b> .....	<b>2</b>
<b>2</b>	<b>Contraindication</b> .....	<b>3</b>
<b>3</b>	<b>Warnings and Precautions</b> .....	<b>3</b>
<b>4</b>	<b>Limitations</b> .....	<b>3</b>
<b>5</b>	<b>Test Principle</b> .....	<b>4</b>
<b>6</b>	<b>FoundationOne Liquid CDx cfDNA Blood Specimen Collection Kit Contents</b> .....	<b>7</b>
<b>7</b>	<b>FoundationOne Liquid CDx Sample Collection and Test Ordering</b> .....	<b>7</b>
<b>8</b>	<b>Instruments</b> .....	<b>7</b>
<b>9</b>	<b>Performance Characteristics</b> .....	<b>7</b>
9.1	Concordance – Comparison to an Orthogonal cfDNA NGS Method #1.....	10
9.2	Concordance – FoundationOne Liquid CDx to validated NGS tumor tissue assay ( <i>BRCA1</i> and <i>BRCA2</i> alterations).....	12
9.3	Concordance – Comparison to an Orthogonal cfDNA NGS Method #2.....	12
9.4	Limit of Detection (Analytical Sensitivity).....	14
9.5	Limit of Blank (LoB).....	16
9.6	Potentially Interfering Substances.....	17
9.7	Hybrid Capture Bait Specificity.....	18
9.8	Carryover/Cross-Contamination.....	18
9.9	Precision: Repeatability and Reproducibility.....	19
9.10	Reagent Lot Interchangeability.....	25
9.11	Variant Curator Precision.....	26
9.12	Stability.....	26
9.12.1	Reagent Stability.....	26
9.12.2	Whole Blood Specimen Stability.....	26
9.13	DNA Extraction.....	29
9.14	Guard Banding/Robustness.....	30
9.15	Pan-Tumor Performance.....	31
9.16	Concordance – FoundationOne Liquid Laboratory Developed Test to FoundationOne Liquid CDx.....	32
9.17	Molecular Index Barcode Performance.....	33
9.18	Automation Line Equivalence.....	33
<b>10</b>	<b>Clinical Validation Studies</b> .....	<b>34</b>
10.1	Clinical Bridging Study: Detection of <i>ALK</i> Rearrangements to Determine Eligibility for Treatment with Alectinib.....	34
10.2	FoundationOne Liquid CDx Concordance Study for <i>EGFR</i> Exon 19 deletion and <i>EGFR</i> Exon 21 L858R Alteration.....	35
10.3	Clinical Bridging Study: Detection of <i>BRCA1/BRCA2/ATM</i> Alterations to Determine Eligibility for Treatment with Olaparib.....	36
10.4	Clinical Bridging Study: Detection of <i>BRCA1</i> and <i>BRCA2</i> Alterations to Determine Eligibility of mCRPC Patients for Treatment with Rucaparib.....	38
10.5	Clinical Validation Study: Detection of <i>BRCA1</i> and <i>BRCA2</i> Alterations to Determine Eligibility of Ovarian Cancer Patients for Treatment with Rucaparib.....	40
10.6	Clinical Bridging Study: Detection of <i>PIK3CA</i> Alterations to Determine Eligibility for Treatment with Alpelisib.....	42
<b>11</b>	<b>CDx Classification Criteria</b> .....	<b>44</b>
11.1	CDx classification criteria for <i>ALK</i> rearrangements, qualifying NSCLC patients for therapy with ALECENSA® (alectinib):.....	44
11.2	CDx classification criteria for <i>EGFR</i> alterations, qualifying NSCLC patients for therapy with IRESSA® (gefitinib), TAGRISSO® (osimertinib), TARCEVA® (erlotinib):.....	44
11.3	CDx classification criteria for <i>BRCA1</i> , <i>BRCA2</i> , and <i>ATM</i> alterations, qualifying prostate cancer patients for therapy with LYNPARZA® (olaparib):.....	44
11.4	CDx classification criteria for <i>BRCA1</i> and <i>BRCA2</i> alterations, qualifying prostate cancer or ovarian cancer patients for therapy with RUBRACA® (rucaparib):.....	45
11.5	CDx classification criteria for <i>PIK3CA</i> alterations, qualifying breast cancer patients for therapy with PIQRAY® (alpelisib):.....	46

## FoundationOne® Liquid CDx Technical Information

Foundation Medicine, Inc.  
150 Second Street, Cambridge, MA 02141  
Phone: 617.418.2200

### 1 Intended Use

FoundationOne Liquid CDx is a qualitative next generation sequencing based *in vitro* diagnostic test that uses targeted high throughput hybridization-based capture technology to detect and report substitutions, insertions and deletions (indels) in 311 genes, rearrangements in four (4) genes and copy number alterations in three (3) genes. FoundationOne Liquid CDx utilizes circulating cell-free DNA (cfDNA) isolated from plasma derived from anti-coagulated peripheral whole blood of cancer patients collected in FoundationOne Liquid CDx cfDNA blood collection tubes included in the FoundationOne Liquid CDx Blood Sample Collection Kit. The test is intended to be used as a companion diagnostic to identify patients who may benefit from treatment with the targeted therapies listed in Table 1 in accordance with the approved therapeutic product labeling.

**Table 1: Companion diagnostic indications**

Tumor Type	Biomarker(s) Detected	Therapy
Non-small cell lung cancer (NSCLC)	<i>ALK</i> rearrangements	ALECENSA® (alectinib)
	<i>EGFR</i> Exon 19 deletions and <i>EGFR</i> Exon 21 L858R substitution	IRESSA® (gefitinib) TAGRISSO® (osimertinib) TARCEVA® (erlotinib)
Prostate cancer	<i>BRCA1</i> , <i>BRCA2</i> , <i>ATM</i> alterations	LYNPARZA® (olaparib)
	<i>BRCA1</i> , <i>BRCA2</i> alterations	RUBRACA® (rucaparib)
Ovarian cancer	<i>BRCA1</i> , <i>BRCA2</i> alterations	RUBRACA® (rucaparib)
Breast cancer	<i>PIK3CA</i> mutations C420R, E542K, E545A, E545D [1635G>T only], E545G, E545K, Q546E, Q546R; and H1047L, H1047R, and H1047Y	PIQRAY® (alpelisib)

Additionally, FoundationOne Liquid CDx is intended to provide tumor mutation profiling to be used by qualified health care professionals in accordance with professional guidelines in oncology for patients with solid malignant neoplasms.

A negative result from a plasma specimen does not mean that the patient's tumor is negative for genomic findings. Patients with the tumor types above who are negative for the mutations listed in **Table 1** should be reflexed to routine biopsy and their tumor mutation status confirmed using an FDA-approved tumor tissue test, if feasible.

Genomic findings other than those listed in Table 1 are not prescriptive or conclusive for labeled use of any specific therapeutic product.

FoundationOne Liquid CDx is a single-site assay performed at Foundation Medicine, Inc. in Cambridge, MA.

## 2 Contraindication

There are no known contraindications.

## 3 Warnings and Precautions

- Alterations reported may include somatic (not inherited) or germline (inherited) alterations; however, the test does not distinguish between germline and somatic alterations. If a reported alteration is suspected to be germline, confirmatory testing should be considered in the appropriate clinical context.
- The test is not intended to replace germline testing or to provide information about cancer predisposition.
- Patients for whom no companion diagnostic alterations are detected should be considered for confirmation with an FDA-approved tumor tissue test, if available.

## 4 Limitations

- For in vitro diagnostic use.
- For prescription use only. This test must be ordered by a qualified medical professional in accordance with clinical laboratory regulations.
- Genomic findings other than those listed in Table 1 of the intended use are not prescriptive or conclusive for labeled use of any specific therapeutic product.
- A negative result does not rule out the presence of an alteration in the patient's tumor.
- Decisions on patient care and treatment must be based on the independent medical judgment of the treating physician, taking into consideration all applicable information concerning the patient's condition, such as patient and family history, physical examinations, information from other diagnostic tests, and patient preferences, in accordance with the standard of care in a given community.
- The test is intended to be performed on specific serial number-controlled instruments by Foundation Medicine, Inc.
- Genomic findings from circulating cell-free DNA (cfDNA) may originate from circulating tumor DNA fragments, germline alterations, or nontumor somatic alterations, such as clonal hematopoiesis of indeterminate potential (CHIP). Genes with alterations that may be derived from CHIP include, but are not limited to, the following: *ASXL1*, *ATM*, *CBL*, *CHEK2*, *DNMT3A*, *JAK2*, *KMT2D (MLL2)*, *MPL*, *MYD88*, *SF3B1*, *TET2*, *TP53*, and *U2AF1*. The efficacy of targeting such nontumor somatic alterations (e.g., CH) is unknown.
- The false positive rate of this test was evaluated in healthy donors. The detection rate for unique short variants in apparently healthy patients is 0.82%. Across 30,622 short variants, 58 variants had a detection rate of greater than 5%.
- The analytical accuracy for the FoundationOne Liquid CDx assay has not been demonstrated in all genes.
- The precision of FoundationOne Liquid CDx was only confirmed for select variants at the limit of detection.
- The FoundationOne Liquid CDx assay does not detect heterozygous deletions.
- The FoundationOne Liquid CDx assay does not detect copy number losses/homozygous deletions in *ATM*.

- A complete assessment of the impact of cfDNA blood collection tube lot-to-lot variability on the performance of the test has not been evaluated.
- The test is not intended to provide information on cancer predisposition.
- *BRCA1/BRCA2* homozygous deletions and rearrangements were not adequately represented in all analytical studies.
- Representation of *ALK* rearrangements were limited in the analytical validation studies.
- The representation of *ATM* short variants and rearrangements was limited in the analytical validation studies.
- Performance has not been validated for cfDNA input below the specified minimum input.

## 5 Test Principle

The FoundationOne Liquid CDx assay is performed exclusively as a laboratory service using circulating cell-free DNA (cfDNA) isolated from plasma derived from anti-coagulated peripheral whole blood from patients with solid malignant neoplasms. The assay employs a single DNA extraction method to obtain cfDNA from plasma from whole blood. Extracted cfDNA undergoes whole-genome shotgun library construction and hybridization-based capture of 324 cancer-related genes. All coding exons of 309 genes are targeted; select intronic or non-coding regions are targeted in fifteen of these genes (refer to **Table 2** for the complete list of genes reported by FoundationOne Liquid CDx). Hybrid-capture selected libraries are sequenced with deep coverage using the NovaSeq® 6000 platform. Sequence data are processed using a custom analysis pipeline designed to detect genomic alterations, including base substitutions and indels in 311 genes, copy number variants in three genes, and genomic rearrangements in four genes. A subset of targeted regions in 75 genes is baited for enhanced sensitivity.

**Table 2: As part of its FDA-approved intended use, the FoundationOne Liquid CDx assay interrogates 324 genes, including 309 genes with complete exonic (coding) coverage and 15 genes with only select non-coding coverage (indicated with an \*).**

Select regions in 75 genes (indicated in bold) are captured with increased sensitivity. Genes are captured for increased sensitivity with complete exonic (coding) coverage unless otherwise noted.

<b>ABL1</b> [Exons 4-9]	<i>ACVR1B</i>	<b>AKT1</b> [Exon 3]	<i>AKT2</i>	<i>AKT3</i>	<b>ALK</b> [Exons 20-29, Introns 18,19]	<i>ALOX12B</i>	<i>AMER1</i> (FAM123B)	<b>APC</b>	<b>AR</b>
<b>ARAF</b> [Exons 4, 5, 7, 11, 13, 15, 16]	<i>ARFRP1</i>	<i>ARID1A</i>	<i>ASXL1</i>	<b>ATM</b>	<b>ATR</b>	<i>ATRX</i>	<i>AURKA</i>	<i>AURKB</i>	<i>AXIN1</i>
<i>AXL</i>	<i>BAP1</i>	<i>BARD1</i>	<i>BCL2</i>	<i>BCL2L1</i>	<i>BCL2L2</i>	<i>BCL6</i>	<i>BCOR</i>	<i>BCORL1</i>	<b>BCR*</b> [Introns 8, 13, 14]
<b>BRAF</b> [Exons 11-18, Introns 7-10]	<b>BRCA1</b> [Introns 2, 7, 8, 12, 16, 19, 20]	<b>BRCA2</b> [Intron 2]	<i>BRD4</i>	<i>BRIP1</i>	<i>BTG1</i>	<i>BTG2</i>	<b>BTK</b> [Exons 2, 15]	<i>C11orf30</i> (EMSY)	<i>C17orf39</i> (GID4)
<i>CALR</i>	<i>CARD11</i>	<i>CASP8</i>	<i>CBFB</i>	<i>CBL</i>	<b>CCND1</b>	<i>CCND2</i>	<i>CCND3</i>	<i>CCNE1</i>	<i>CD22</i>
<i>CD70</i>	<b>CD74*</b> [Introns 6-8]	<i>CD79A</i>	<i>CD79B</i>	<b>CD274</b> (PD-L1)	<i>CDC73</i>	<b>CDH1</b>	<b>CDK12</b>	<b>CDK4</b>	<b>CDK6</b>

CDK8	CDKN1A	CDKN1B	<b>CDKN2A</b>	CDKN2B	CDKN2C	CEBPA	CHEK1	<b>CHEK2</b>	CIC
CREBBP	<b>CRKL</b>	CSF1R	CSF3R	CTCF	CTNNA1	<b>CTNNB1</b> [Exon 3]	CUL3	CUL4A	CXCR4
CYP17A1	DAXX	DDR1	<b>DDR2</b> [Exons 5, 17, 18]	DIS3	DNMT3A	DOT1L	EED	<b>EGFR</b> [Introns 7, 15, 24-27]	EP300
EPHA3	EPHB1	EPHB4	<b>ERBB2</b>	<b>ERBB3</b> [Exons 3, 6, 7, 8, 10, 12, 20, 21, 23, 24, 25]	ERBB4	ERCC4	ERG	<b>ERRFI1</b>	<b>ESR1</b> [Exons 4-8]
<i>ETV4*</i> [Intron 8]	<i>ETV5*</i> [Introns 6, 7]	<b>ETV6*</b> [Introns 5, 6]	<i>EWSR1*</i> [Introns 7-13]	<b>EZH2</b> [Exons 4, 16, 17, 18]	<i>EZR*</i> [Introns 9-11]	FAM46C	FANCA	FANCC	FANCG
FANCL	FAS	FBXW7	FGF10	FGF12	FGF14	FGF19	FGF23	FGF3	FGF4
FGF6	<b>FGFR1</b> [Introns 1, 5, Intron 17]	<b>FGFR2</b> [Intron 1, Intron 17]	<b>FGFR3</b> [Exons 7, 9 (alternative designation exon 10), 14, 18, Intron 17]	FGFR4	FH	FLCN	FLT1	<b>FLT3</b> [Exons 14, 15, 20]	<b>FOXL2</b>
FUBP1	GABRA6	GATA3	GATA4	GATA6	<b>GNA11</b> [Exons 4, 5]	GNA13	<b>GNAQ</b> [Exons 4, 5]	<b>GNAS</b> [Exons 1, 8]	GRM3
GSK3B	H3F3A	HDAC1	HGF	HNF1A	<b>HRAS</b> [Exons 2, 3]	HSD3B1	ID3	<b>IDH1</b> [Exon 4]	<b>IDH2</b> [Exon 4]
IGF1R	IKBKE	IKZF1	INPP4B	IRF2	IRF4	IRS2	JAK1	<b>JAK2</b> [Exon 14]	<b>JAK3</b> [Exons 5, 11, 12, 13, 15, 16]
JUN	KDM5A	KDM5C	KDM6A	KDR	KEAP1	KEL	<b>KIT</b> [Exons 8, 9, 11, 12, 13, 17, Intron 16]	KLHL6	<b>KMT2A</b> (MLL) [Introns 6, 8-11, Intron 7]
<b>KMT2D</b> (MLL2)	<b>KRAS</b>	LTK	LYN	MAF	<b>MAP2K1</b> (MEK1) [Exons 2, 3]	<b>MAP2K2</b> (MEK2) [Exons 2-4, 6, 7]	MAP2K4	MAP3K1	MAP3K13
MAPK1	MCL1	<b>MDM2</b>	MDM4	MED12	MEF2B	MEN1	MERTK	<b>MET</b>	MITF
MKNK1	MLH1	<b>MPL</b> [Exon 10]	MRE11A	<i>MSH2</i> [Intron 5]	MSH3	MSH6	MST1R	MTAP	<b>MTOR</b> [Exons 19, 30, 39, 40, 43-45, 47, 48, 53, 56]
MUTYH	<i>MYB*</i> [Intron 14]	<b>MYC</b> [Intron 1]	<i>MYCL</i> (MYCL1)	<b>MYCN</b>	<b>MYD88</b> [Exon 4]	NBN	<b>NF1</b>	NF2	NFE2L2
NFKBIA	<i>NKX2-1</i> (TTF-1)	NOTCH1	<i>NOTCH2</i> [Intron 26]	NOTCH3	<b>NPM1</b> [Exons 4-6, 8, 10]	<b>NRAS</b> [Exons 2, 3]	<i>NSD3</i> (WHSC1L1)	NT5C2	<b>NTRK1</b> [Exons 14, 15, Introns 8-11]

<i>NTRK2</i> [Intron 12]	<b>NTRK3</b> [Exons 16, 17]	<i>NUTM1*</i> [Intron 1]	<i>P2RY8</i>	<b>PALB2</b>	<i>PARK2</i>	<i>PARP1</i>	<i>PARP2</i>	<i>PARP3</i>	<i>PAX5</i>
<i>PBRM1</i>	<i>PDCD1</i> (PD-1)	<b>PDCD1LG2</b> (PD-L2)	<b>PDGFRA</b> [Exons 12, 18, Introns 7, 9, 11]	<b>PDGFRB</b> [Exons 12-21, 23]	<i>PDK1</i>	<i>PIK3C2B</i>	<i>PIK3C2G</i>	<b>PIK3CA</b> [Exons 2, 3, 5-8, 10, 14, 19, 21 (Coding Exons 1, 2, 4-7, 9, 13, 18, 20)]	<i>PIK3CB</i>
<i>PIK3R1</i>	<i>PIM1</i>	<i>PMS2</i>	<i>POLD1</i>	<i>POLE</i>	<i>PPARG</i>	<i>PPP2R1A</i>	<i>PPP2R2A</i>	<i>PRDM1</i>	<i>PRKAR1A</i>
<i>PRKCI</i>	<i>PTCH1</i>	<b>PTEN</b>	<b>PTPN11</b>	<i>PTPRO</i>	<i>QKI</i>	<i>RAC1</i>	<i>RAD21</i>	<i>RAD51</i>	<i>RAD51B</i>
<i>RAD51C</i>	<i>RAD51D</i>	<i>RAD52</i>	<i>RAD54L</i>	<b>RAF1</b> [Exons 3, 4, 6, 7, 10, 14, 15, 17, Introns 4-8]	<i>RARA</i> [Intron 2]	<b>RB1</b>	<i>RBM10</i>	<i>REL</i>	<b>RET</b> [Introns 7, 8, Exons 11, 13-16, Introns 9-11]
<i>RICTOR</i>	<i>RNF43</i>	<b>ROS1</b> [Exons 31, 36-38, 40, Introns 31-35]	<i>RPTOR</i>	<i>RSPO2*</i> [Intron 1]	<i>SDC4*</i> [Intron 2]	<i>SDHA</i>	<i>SDHB</i>	<i>SDHC</i>	<i>SDHD</i>
<i>SETD2</i>	<i>SF3B1</i>	<i>SGK1</i>	<i>SLC34A2*</i> [Intron 4]	<i>SMAD2</i>	<i>SMAD4</i>	<i>SMARCA4</i>	<i>SMARCB1</i>	<b>SMO</b>	<i>SNCAIP</i>
<i>SOCS1</i>	<i>SOX2</i>	<i>SOX9</i>	<i>SPEN</i>	<i>SPOP</i>	<i>SRC</i>	<i>STAG2</i>	<i>STAT3</i>	<b>STK11 (LKB1)</b>	<i>SUFU</i>
<i>SYK</i>	<i>TBX3</i>	<i>TEK</i>	<i>TERC*</i> {ncRNA}	<b>TERT*</b> {Promoter}	<i>TET2</i>	<i>TGFBR2</i>	<i>TIPARP</i>	<i>TMPRSS2*</i> [Introns 1-3]	<i>TNFAIP3</i>
<i>TNFRSF14</i>	<b>TP53</b>	<i>TSC1</i>	<i>TSC2</i>	<i>TYRO3</i>	<i>U2AF1</i>	<b>VEGFA</b>	<i>VHL</i>	<i>WHSC1</i>	<i>WT1</i>
<i>XPO1</i>	<i>XRCC2</i>	<i>ZNF217</i>	<i>ZNF703</i>						

The classification criteria for all CDx variants are outlined at the end of this document.

The output of the test includes:

Category 1: Companion Diagnostic (CDx) claims noted in **Table 1** of the Intended Use

Category 2: cfDNA Biomarkers with Strong Evidence of Clinical Significance in cfDNA

Category 3: Biomarkers with Evidence of Clinical Significance in tissue supported by:

3A: strong analytical validation using cfDNA

3B: analytical validation using cfDNA

Category 4: Other Biomarkers with Potential Clinical Significance

As part of its FDA-approved intended use, copy number alterations and rearrangements are reported in the genes listed in **Table 3**.

**Table 3: Genes for which copy number alterations and rearrangements are reported for tumor profiling by FoundationOne Liquid CDx**

Alteration Type	Genes
Copy Number Alterations	<i>BRCA1, BRCA2, ERBB2</i>
Rearrangements	<i>ALK, BRCA1, BRCA2</i>

## 6 FoundationOne Liquid CDx cfDNA Blood Specimen Collection Kit Contents

### Test Kit Contents

The test includes a sample shipping kit, which is sent to ordering laboratories. The shipping kit contains the following components:

- Specimen preparation and shipping instructions
- Two FoundationOne Liquid CDx cfDNA blood collection tubes (8.5 mL nominal fill volume per tube)
- Return shipping label

All other reagents, materials and equipment needed to perform the assay are used exclusively in the Foundation Medicine laboratory. The FoundationOne Liquid CDx assay is intended to be performed with serial number-controlled instruments.

## 7 FoundationOne Liquid CDx Sample Collection and Test Ordering

To order FoundationOne Liquid CDx, the test order form in the test kit must be fully completed and signed by the ordering physician or other authorized medical professional. Please refer to Specimen Preparation Instructions and Shipping Instructions included in the test kit.

For more detailed information, including Performance Characteristics, please find the FDA Summary of Safety and Effectiveness Data at: [https://www.accessdata.fda.gov/cdrh\\_docs/pdf19/P190032B.pdf](https://www.accessdata.fda.gov/cdrh_docs/pdf19/P190032B.pdf)

## 8 Instruments

The FoundationOne Liquid CDx device is intended to be performed with the following instruments, as identified by specific serial numbers:

- Illumina NovaSeq 6000
- Beckman Biomek NXP Span-8 Liquid Handler
- Thermo Scientific Kingfisher Flex DW 96
- Bravo Benchbot
- Hamilton STARTlet-STAR Liquid Handling Workstation

## 9 Performance Characteristics

Performance characteristics were established using contrived and clinical circulating cfDNA derived from blood specimens extracted from a wide range of tumor types. **Table 4** below provides a summary of the number of tumor types and variants included in each study. As summarized in this table, each study included a broad range of representative alteration types (substitutions, insertion-deletions, copy number alterations, rearrangements) in various genomic contexts across a number of genes. The validation studies included >7,000 sample replicates, >31,000 unique variants [includes variants classified as variants of unknown significance (VUS) and/or benign], >30 tumor types, representing all 324 genes targeted by the assay.

**Table 4 Representation of tumor types and variants across validation studies**

Study Title	Cancer Types Represented	# Unique Samples	# of Sample Replicates	# of Unique Genes	# of Unique				
					Subs	Indels	Rearrang.	Copy Number Amplif.	Copy Number Losses
Contrived Sample Functional Characterization (CSFC) Study	Breast cancer Colorectal cancer Lung cancer Contrived samples	13	1843	228	563	81	11	1	1
FoundationOne Liquid CDx to Validated NGS Tumor Tissue Test Concordance: <i>BRCA1</i> and <i>BRCA2</i> Variants	Prostate cancer Ovarian cancer	279	N/A	2	100	87	9	0	2
FoundationOne Liquid CDx to Validated NGS cfDNA Assay Concordance: <i>PIK3CA</i> mutations	Breast cancer	412	N/A	1	32	5	0	0	0
Orthogonal Concordance	23 cancer types Contrived samples	278	N/A	64	541	12	11	3	0
LoD Estimation	Prostate Contrived samples	10	877	286	1490	247	32	13	3
LoB	Healthy Donors	28	79	322	26134	4482	911	222	42
Potentially Interfering Substances	Contrived samples	9	336	18	16	11	11	1	2
Hybrid Capture Bait Specificity	25 cancer types Contrived samples	3546	N/A	324	N/A	N/A	N/A	N/A	N/A
Reagent Stability	Contrived samples	8	142	279	1090	215	32	17	2
Reagent Interchangeability	Contrived samples	8	192	20	15	11	11	1	1
Precision study 1	Breast cancer Colon cancer Lung cancer Ovarian cancer Prostate cancer Skin cancer Contrived samples	47	1121	280	900	229	63	49	5
Precision study 2	Lung cancer Prostate cancer Stomach cancer Colorectal cancer Bile duct cancer Breast cancer	10	230	6	6	4	0	0	0
DNA Extraction	Colorectal cancer Prostate cancer Breast cancer Lung cancer Skin cancer	6	72	161	265	53	2	0	0
Whole Blood Sample Stability	Lung cancer Colorectal cancer Prostate cancer Breast cancer	11	22	66	75	15	1	0	0



Study Title	Cancer Types Represented	# Unique Samples	# of Sample Replicates	# of Unique Genes	# of Unique				
					Subs	Indels	Rearrang.	Copy Number Amplif.	Copy Number Losses
Inverted Tube Whole Blood Sample Stability	Lung cancer Colorectal cancer Breast cancer Ovarian cancer Prostate cancer	130	260	237	594	91	5	5	0
Cross Contamination	Contrived samples	5	376	39	9	5	4	21	1
Guard Banding	Contrived samples	10	375	20	17	12	12	1	1
Clinical validation for detection of <i>EGFR</i> exon 19 deletions and L858R alterations: non-inferiority study	Lung cancer	177	N/A	1	5	7	N/A	N/A	N/A
Clinical validation study for detection of deleterious alterations in <i>BRCA1</i> and <i>BRCA2</i> in prostate cancer	Prostate cancer	199	N/A	2	44	55	8	0	1
Clinical validation study for detection of deleterious alterations in <i>BRCA1</i> and <i>BRCA2</i> in ovarian cancer	Ovarian cancer	217	N/A	2	48	49	3	0	0
Clinical validation study for detection of PIK3CA mutations in breast cancer	Breast	359	N/A	1	28	4	0	0	0
Clinical validation study for ALK rearrangements in NSCLC	Lung cancer	249	N/A	1	13	1	11	1	0
Clinical validation study for <i>BRCA1</i> , <i>BRCA2</i> , and <i>ATM</i> alterations in prostate cancer	Prostate cancer	333	N/A	3	48	75	10	0	1

Study Title	Cancer Types Represented	# Unique Samples	# of Sample Replicates	# of Unique Genes	# of Unique				
					Subs	Indels	Rearrang.	Copy Number Amplif.	Copy Number Losses
Blood Collection Tube Equivalence	Ovarian cancer Breast cancer Colorectal cancer Prostate cancer Lung cancer Skin cancer Stomach cancer	60	192	116	135	39	13	5	0
Automation Line Equivalence	Contrived samples	8	187	303	1926	337	63	61	4
Variant Report Curation	Breast cancer Colorectal cancer Lung cancer Prostate cancer Skin cancer	19	57	183	300	104	15	11	2
Pan-tumor performance (includes historical analysis)	20 cancer types	19868	N/A	N/A	N/A	N/A	N/A	N/A	N/A
Molecular Index Barcode Performance	25 cancer types Contrived samples	7637	N/A	324	N/A	N/A	N/A	N/A	N/A
FoundationOne Liquid LDT to FoundationOne Liquid CDx Concordance	25 cancer types	927	N/A	73	1815	376	109	46	N/A

\* Variants detected include variants classified as VUS and benign.

### 9.1 Concordance – Comparison to an Orthogonal cfDNA NGS Method #1

The detection of short variants and rearrangements by the FoundationOne Liquid CDx assay was compared to that of an externally validated NGS assay in 74 genes common to both assays across 278 samples that represented an array of tumor types (>50 unique disease ontologies across 23 cancer types). The cancer types (# samples) included lung [NSCLC (75) and other (3)]; breast (54); prostate (32); colorectal [colon (27) and rectal (6)]; liver (11); ovarian (6); pancreas (9); gastrointestinal (7); bile duct (2); esophageal (5); skin (6); cervical (1); anal (1); bladder (1); gallbladder (1); salivary gland (2); thymus (1); thyroid (3); uterine (2); fallopian tube (1); head and neck (1); soft tissue (1); and unknown primary (19). The study included samples selected from clinical FoundationOne Liquid testing (n=268) and contrived samples consisting of fragmented gDNA diluted in clinical cfDNA to represent rare alterations (n=10).

Using the externally validated NGS assay as the comparator, the analysis demonstrated a short variant PPA of 96.2% with a 95% two-sided CI of [94.8%-97.4%]. The short variant NPA was >99.9% with a 95% two-sided CI of [99.9%-100.0%]. The respective PPA of base substitutions and indels with a 95% two-sided CI was 96.1%

[94.6%-97.3%] and 100.0% [85.2%-100.0%]. The respective NPA and 95% two-sided CI of base substitutions and indels was >99.9% [99.9%-100.0%] and 100.0% [99.89%-100.0%] (Table 5).

**Table 5. Concordance of short variants called in FoundationOne Liquid CDx and the comparator assay (n= 902 positive variants, n= 152,832 negative variants\* by the comparator assay)**

Variant Type	FoundationOne Liquid CDx(+) Comparator(+)	FoundationOne Liquid CDx(-) Comparator(+)	FoundationOne Liquid CDx(+) Comparator(-)	FoundationOne Liquid CDx(-) Comparator(-)	PPA [95% CI]	NPA [95% CI]	OPA [95% CI]
All Short Variants	868	34	8	152824	96.2% [94.8%-97.4%]	>99.9% [99.9%-100.0%]	>99.9% [99.9%-100.0%]
Base Substitutions	845	34	8	149511	96.1% [94.6%-97.3%]	>99.9% [99.9%-100.0%]	>99.9% [99.9%-100.0%]
Indels	23	0	0	3361	100.0% [85.2%- 100.0%]	100.0% [99.9%- 100.0%]	100.0% [99.9%- 100.0%]

\* Variants detected include variants classified as VUS and benign.

For the concordance of rearrangement detection between FoundationOne Liquid CDx and the comparator assay, the observed rearrangement PPA was 100.0%, with a 95% two-sided CI of [59.0%-100.0%]. The NPA was 99.8%, with a 95% two-sided CI [99.5%-100.0%] (Table 6).

**Table 6. Concordance of rearrangements called in FoundationOne Liquid CDx and the comparator assay (n= 7 positive, n=1685 negative\* as determined by the comparator assay)**

	Comparator (+)	Comparator (-)	Total
FoundationOne Liquid CDx (+)	7	3	10
FoundationOne Liquid CDx (-)	0	1682	1682
Total	7	1685	1692
	PPA: 100.0% [59.0% - 100.0%]	NPA: 99.8% [99.5% - 100.0%]	OPA: 99.8% [99.5% - 100.0%]

\* Variants detected include variants classified as VUS and benign.

Assessment of a subset of highly-actionable alterations were compared between the two assays. The analysis resulted in a PPA of 100% across all eligible highly-actionable alterations called in the comparator assay (Table 7).

**Table 7. Concordance of CDx alterations called between FoundationOne Liquid CDx and the comparator assay (n = 74)**

Targeted Alteration	n	PPA [95% CI]	NPA [95% CI]
BRCA1 short variants	1	100% [2.5%-100.0%]	100% [98.7%-100.0%]
BRCA2 short variants	2	100% [15.8%-100.0%]	100% [99.3%-100.0%]
EGFR exon 19 deletions	11	100% [71.5%-100.0%]	100% [99.7%-100.0%]
EGFR L858R	10	100% [69.2%-100.0%]	100% [98.7%-100.0%]
PIK3CA base substitutions	49	100% [92.7%-100.0%]	100% [99.9%-100.0%]

Targeted Alteration	n	PPA [95% CI]	NPA [95% CI]
ALK rearrangements	1	100% [2.5%-100.0%]	99.9% [99.7%-100.0%]

These data demonstrate that the FoundationOne Liquid CDx assay and an externally-validated NGS assay are highly concordant across the 74 genes common between the two panels.

### 9.2 Concordance – FoundationOne Liquid CDx to validated NGS tumor tissue assay (*BRCA1* and *BRCA2* alterations)

Samples from a total of 279 prostate and ovarian cancer patients were tested and the concordance evaluated between FoundationOne Liquid CDx and the validated NGS tumor tissue assay for the detection of deleterious alterations in *BRCA1* or *BRCA2*. As summarized below, a PPA of 88.03% and an NPA of 95.68% were observed on a sample level (**Table 8**). As summarized in **Table 9**, an overall PPA of 87.28% and an NPA of 99.83% were observed at the variant level. Some discordance is expected based on biological differences and sampling times between tumor tissue and plasma samples. Considering the impact of biological differences between analytes, these data demonstrate a high concordance between FoundationOne Liquid CDx and the validated NGS tumor tissue assay for the detection of deleterious alterations in *BRCA1* or *BRCA2*.

**Table 8. Concordance (by sample) of FoundationOne Liquid CDx and validated NGS tumor tissue assay in prostate and ovarian cancer patients for the detection of alterations in *BRCA1* or *BRCA2***

		NGS Tumor Tissue Assay	
		Positive	Negative
FoundationOne Liquid CDx	Positive	103	7
	Negative	14	155
		PPA: 88.03% [80.91%-92.74%]	NPA: 95.68% [91.35%-97.89%]

**Table 9. Concordance (by variant) of FoundationOne Liquid CDx and validated NGS tumor tissue assay in prostate and ovarian cancer patients for the detection of alterations in *BRCA1* or *BRCA2***

	F1LCDx+ /Tissue+	F1L CDx- /Tissue+	F1L CDx+ /Tissue-	F1L CDx-/ Tissue-	PPA (95% CI)	NPA (95% CI)
Substitutions	77	6	29	20255	92.77% (85.11%, 96.64%)	99.86% (99.79%, 99.90%)
Indels	65	3	31	16362	95.59% (87.81%, 98.49%)	99.81% (99.73%, 99.87%)
Rearrangements	4	3	7	1939	57.14% (25.05%, 84.18%)	99.64% (99.26%, 99.83%)
Copy number loss	5	10	1	263	33.33% (15.18%, 58.29%)	99.62% (97.89%, 99.93%)
Total	151	22	68	38819	87.28% (81.50%, 91.45%)	99.83% (99.78%, 99.86%)

### 9.3 Concordance – Comparison to an Orthogonal cfDNA NGS Method #2

The accuracy of using FoundationOne Liquid CDx as a companion diagnostic to identify breast cancer patients harboring *PIK3CA* alterations was assessed with residual plasma samples from the SOLAR-1 clinical trial. Of the remaining plasma samples, 542 were evaluable by the externally-validated NGS method and produced valid results. 418 were evaluable by FoundationOne Liquid CDx, of which 192 positive variants were detected across 188 patients, with four patients possessing two positive variants each. The distribution of counts per positive variant is listed in **Table 10**.

**Table 10. Distribution of variants detected with FoundationOne Liquid CDx evaluable samples.**

Protein Effect in PIK3CA	# Variant Calls (188 Positive Samples)
C420R	3
E542K	25
E545A	1
E545G	2
E545K	50
H1047L	9
H1047R	100
H1047Y	1
Q546R	1
<b>Total</b>	<b>192</b>

A total of 412 valid samples generated valid results with both assays. The primary analysis using NGS Method #2 as the reference assay achieved a PPA [95% CI] of 97.06% [93.27%, 99.04%], and an NPA [95% CI] of 91.74% [87.52%, 94.88%]. The contingency table for this comparison is provided in **Table 11**, with counts representing number of samples (versus number of variant calls).

The sample counts in the core 2x2 white boxes total to 412 samples. There were seven samples evaluable with FoundationOne Liquid CDx but failed (italicized in **Table 11**), as well as three samples missing from reference assay data. There were five samples unevaluable by the reference assay; three of these aligned with the 418 evaluable FoundationOne Liquid CDx samples, while two were among the 130 samples not evaluable due to insufficient plasma.

**Table 11. Contingency table comparing FoundationOne Liquid CDx with the reference assay, primary analysis with 412 cases.**

		Reference Assay					
		Positive	Negative	Not Evaluable	Missing	Total	
FoundationOne Liquid CDx	<b>Positive</b>	165	20	2	1	<b>188</b>	PPA <sub>F1L</sub> : 89.19% [83.80%, 93.27%]
	<b>Negative</b>	5	222	1	2	<b>230</b>	NPA <sub>F1L</sub> : 97.80% [94.93%-99.28%]
	<b>Evaluable but Failed</b>	0	7	0	0	7	
	<b>Not Evaluable</b>	35	93	2	0	130	
	<b>Total</b>	205	342	5	3	555	
		PPA <sub>ONC</sub> : 97.06% [93.27%, 99.04%]	NPA <sub>ONC</sub> : 91.74% [87.52%, 94.88%]				OPA: 93.93% [91.17%, 96.04%]

#### 9.4 Limit of Detection (Analytical Sensitivity)

The LoD for each variant type was established by processing a total of 1,069 sample replicates across ten contrived (enzymatically fragmented cell-line gDNA) samples representing short variants, rearrangements, and copy number alterations. The LoD was determined using the conservative hit rate approach for the majority of variants. A probit model was used when appropriate (when  $\geq 3$  dilution levels with hit rates between 10% and 90% were observed). LoD by hit rate was defined as the mean VAF value (for short variants and rearrangements) or mean tumor fraction value (for copy number alterations) at the lowest dilution level tested with at least 95% detection across replicates. The hit rate was computed as the number of replicates with positive variant calls per the total number of replicates tested at each level of the targeted VAF (short variants and rearrangements) or tumor fraction (copy number alterations). Short variants with hit rates of at least 95% at all dilution levels or hit rates below 95% for all dilution levels were excluded from analysis as LoD could not be reliably estimated.

The median estimated LoD for CDx alterations are presented in **Table 12**. The median LoD for targeted short variant, rearrangement, and copy number alterations were consistent with the platform LoD (**Table 13**).

**Table 12: LoD estimation for CDx alterations**

Gene	Alteration Subtype	Number of Samples Evaluated	Median LoD*
ATM	Indels	1	0.51% VAF
	Rearrangement ( <i>ATM-EXPH5</i> Truncation <sup>1</sup> )	1	1.13% VAF
BRCA1	Substitutions	8	0.34% VAF
	Indels	1	0.38% VAF*
	Rearrangement <sup>1</sup>	1	0.87% VAF
BRCA2	Substitutions	17	0.37% VAF
	Indels	2	0.36% VAF
	<i>BRCA2- EDA</i> Truncation <sup>1</sup>	1	0.48% VAF
	Copy Number Loss <sup>1</sup>	1	48.1% TF
EGFR	Substitutions (L858R substitutions)	2	0.34% VAF
	Indels (exon 19 deletions)	2	0.27% VAF
PIK3CA	Substitutions	6	0.34% VAF
ALK	Rearrangement ( <i>ALK-EML4</i> )	1	0.24% VAF
	Rearrangement ( <i>NPM1-ALK</i> Rearrangement)	1	0.94% VAF

The estimated LoDs for *BRCA1* and *BRCA2* subs and indels were confirmed at values higher than the LoDs established in **Table 12** (see Precision: Reproducibility and Reproducibility section below, **Table 21** and **Table 22** for confirmed LoD values).

<sup>1</sup>The LoD for these alterations was determined using clinical specimens.

\*The accuracy of %VAF/%TF have not been analytically validated

The platform LoD for short variants, rearrangements, and copy number losses are presented in **Table 13**. A total of 864 short variants were included in the platform LoD analysis. The enhanced sensitivity region of the bait set contains 269 of the short variants analyzed and the standard sensitivity region of the bait set contains 595 of the short variants analyzed. The estimated LoD for short variants is 0.40% for the enhanced sensitivity region and 0.82% of the standard sensitivity region. The median LoD is 30.4% tumor fraction for copy number losses.

Because a major component driving the detectability of a variant is genomic context (repetitiveness of the reference genomic region), the LoD analysis for short variants was also evaluated within categories based on genomic context as summarized in **Table 14**.

**Table 13: LoD estimation by variant type**

Alteration Type	Number of Variants in Analysis	Bait Set Region	Median LoD*	Quartile 1 to Quartile 3 LoD Range
Short Variants	269	Enhanced Sensitivity	0.40% VAF	0.33% - 0.50% VAF
	595	Standard Sensitivity	0.82% VAF	0.70% - 0.98% VAF
Rearrangements	7	Enhanced Sensitivity	0.37% VAF	0.26% - 0.47% VAF
	1	Standard Sensitivity	0.90% VAF	N/A
Copy Number Amplifications	8	N/A	21.7% TF	19.8% - 25.2% TF

VAF = variant allele frequency

TF = tumor fraction

\*The accuracy of %VAF/%TF have not been analytically validated

**Table 14: LoD by variant subtype based on genomic context**

Region	Alteration Subtype	N	Minimum LoD (VAF/TF) <sub>1</sub> *	1st Quartile LoD (VAF/TF) <sub>1</sub>	Median LoD (VAF/TF) <sub>1</sub>	3rd Quartile LoD (VAF/TF) <sub>1</sub>
Enhanced Sensitivity Region	Short Variants: Enhanced Sensitivity Region Total	269	0.20%	0.33%	0.40%	0.50%
	Insertion/Deletion in non-repetitive region or a repetitive region of <=3 base pairs	10	0.23%	0.29%	0.31%	0.36%
	Insertion/Deletion in a repetitive region of 4 to 6 base pairs	23	0.28%	0.37%	0.48%	0.56%
	Insertion/Deletion in a repetitive region of >=7 base pairs	6	0.33%	0.48%	0.58%	0.82%
	Substitution in a non-repetitive region or a repetitive region of <=7 base pairs	229	0.20%	0.33%	0.39%	0.49%
	Substitution in a repetitive region of >7 base pairs	1	0.32%	0.32%	0.32%	0.32%
High Sensitivity Region	Short Variants: High Sensitivity Region Total	595	0.40%	0.70%	0.82%	0.98%
	Insertion/Deletion in non-repetitive region or a repetitive region of <=3 base pairs	18	0.46%	0.68%	0.87%	1.00%
	Insertion/Deletion in a repetitive region of 4 to 6 base pairs	32	0.61%	0.75%	0.87%	0.95%

Standard Sensitivity Region	Insertion/Deletion in a repetitive region of $\geq 7$ base pairs	11	0.59%	1.07%	1.15%	1.20%
	Substitution in a non-repetitive region or a repetitive region of $\leq 7$ base pairs	524	0.40%	0.70%	0.81%	0.96%
	Substitution in a repetitive region of $> 7$ base pairs	8	0.69%	0.83%	0.96%	1.28%
Enhanced Sensitivity Region	Rearrangements	7	0.20%	0.26%	0.37%	0.47%
Enhanced/Standard Sensitivity Region	Rearrangements	1	0.28%	0.28%	0.28%	0.28%
Standard Sensitivity Region	Rearrangements	1	0.90%	0.90%	0.90%	0.90%
NA	Copy Number Amplifications	8	19.8%	19.8%	21.7%	25.2%

1VAF reported for short variant and rearrangement LoD, tumor fraction reported for copy number alterations LoD.

\*The accuracy of %VAF/%TF have not been analytically validated

The median LoD for highly-actionable, non-CDx alterations evaluated for LoD are presented in **Table 15**. The median LoD for these targeted short variants are consistent with the platform LoD presented in **Table 13**.

**Table 15: LoD for non-CDx alterations**

Gene	Alteration Subtype	Number of Samples Evaluated	Median LoD*
<i>BRAF</i>	Substitutions	1	0.33% VAF
<i>KRAS</i>	Substitutions	2	0.33% VAF
<i>MET</i>	Indels	1	0.41% VAF
<i>NRAS</i>	Substitutions	2	0.42% VAF
<i>PALB2</i>	Indels	1	0.37% VAF
	Substitutions	1	0.51% VAF
<i>ERBB2</i>	Copy Number Amplification	1	19.8% TF

VAF = variant allele frequency

TF = tumor fraction

<sup>1</sup> LoD for these alterations was determined using clinical specimens.

\*The accuracy of %VAF/%TF have not been analytically validated

## 9.5 Limit of Blank (LoB)

Per CLSI EP17-A2, the limit of blank (LoB) was established by profiling plasma samples from 30 asymptomatic donors with no diagnosis of cancer with 4 replicates per sample. All donors were over the age of 60 with a median age of 68 and included 15 smokers and 15 non-smokers.

As would be expected in a sampling of human plasma, especially plasma from an aged population, a small number of alterations were detected. Across 30,622 short variants, which include variants classified as VUS/benign, five variants of unknown significance had a detection rate significantly exceeding 5% on an individual variant basis:



*TSC1* 965T>C, *IRF4* 1ins87, *MSH3* 186\_187insGCCGCAGCGCCCGCAGCG, *IGF1R* 568C>T, *WHSC1* 1582C>A.

All other variants were determined to have an LoB of 0, based on the detection rate not significantly exceeding 5%. Each cancer-related alteration detected in this study was detected in replicates from a single donor, indicating that these are likely true variants present in the sample. On a per variant basis (number of unique variants detected at least once across all replicates divided by the total number of unique variants included in the analysis), the overall detection rate for short variants in this study was 0.82%. On a per variant basis (number of variants detected across all replicates divided by the total number of variants included in the analysis across all replicates), the overall detection rate for short variants in this study was 0.027% (**Table 16**).

**Table 16: Detection rate for each reporting category in LoB study**

Category	Unique Variant Detection Rate (Unique variants detected) / (total unique variants analyzed)	Total Variant Detection Rate (Total variants detected) / (total variants analyzed <sup>1</sup> )
Level 1	0% (0 of 292)	0% (0 of 23,068)
Level 2	0% (0 of 10)	0% (0 of 790)
Level 3	0% (0 of 18)	0% (0 of 1,422)
Level 4	0.82% (47 of 5,760)	0.024% (107 of 455,040)
VUS	0.83% (203 of 24,542)	0.029% (555 of 1,938,818)
All categories	0.82% (250 of 30,622)	0.027% (662 of 2,419,138 <sup>1</sup> )

<sup>1</sup> total variants analyzed = unique variants \* 79 replicates

Across 264 copy number alterations and 894 rearrangements, zero variants were detected. These results demonstrate the high specificity of FoundationOne Liquid CDx.

## 9.6 Potentially Interfering Substances

To evaluate the robustness of the FoundationOne Liquid CDx results in the presence of potentially interfering exogenous and endogenous substances, a total of 11 potential interferents were evaluated. These potential interferents included six endogenous substances (albumin, conjugated bilirubin, unconjugated bilirubin, cholesterol, hemoglobin and triglycerides) and five exogenous substances (DNA from another source [the microorganism *Staphylococcus epidermidis*], excess anticoagulant, proteinase K, ethanol and molecular index barcodes).

A total of 340 samples were tested to evaluate the potential interference of albumin, conjugated bilirubin, unconjugated bilirubin, cholesterol, hemoglobin, triglycerides, DNA from another source (the microorganism *Staphylococcus epidermidis*), excess anticoagulant, proteinase K, ethanol, and molecular index barcodes. An assessment of the cfDNA yield obtained during the DNA isolation, purification, and quantification steps, as well as at library construction QC (LCQC) and hybrid capture QC (HCQC) was performed. The process success rates for each step are listed in **Table 17**.

**Table 17 Process success rates with interfering substances**

Process	# Failed	# Pass	Total	Success Rate (%)	95% CI LB (%)	95% CI UB (%)
DNA Extraction	0	180	180	100.00	97.97	100.00
LC	1	339	340	99.71	98.37	99.99
HC	3	336	339	99.12	97.44	99.82
Sequencing	0	336	336	100.00	98.91	100.00

For each potential interferent, concordance of alteration calls was calculated relative to a control sample without interferent. The pre-defined variants included 27 short variants, 17 rearrangements, and 3 copy number variants. Of the 11 potential interferents tested across 16 conditions, concordance for all variant calls was 100% for 8 conditions and ≥97% for all conditions (**Table 18**).

**Table 18: Concordance per substance for variants  $\geq 1x$  LoD**

Substance	Concordance	95% CI LB (Exact)	95% CI UB (Exact)	N
Triglycerides, 37 mmol/L (or 33 g/L)	100.00%	91.19%	100.00%	40
Hemoglobin, 2.0 g/L	100.00%	90.97%	100.00%	39
Albumin, 60 g/L	97.56%	87.14%	99.94%	41
Bilirubin (conjugated), 0.2 g/L	100.00%	91.59%	100.00%	42
Bilirubin (unconjugated), 0.2 g/L	97.44%	86.52 %	99.94%	39
Cholesterol Level 2, 3.88 mmol (150 mg/dL)	97.56%	87.14%	99.94%	41
Cholesterol Level 1, 6.47mmol (250 mg/dL)	97.37%	86.19%	99.93%	38
Staphylococcus epidermidis, 1 x 10 <sup>6</sup> CFU/mL	100.00%	90.97%	100.00%	39
Anticoagulant, 5X nominal volume	100.00%	91.40%	100.00%	41
Proteinase K, +0.6 mg/mL	98.00%	89.35%	99.95%	50
Proteinase K, +0.3 mg/mL	100.00%	92.29%	100.00%	46
Ethanol, +2.5%	97.96%	89.15%	99.95%	49
Ethanol, +5.0%	97.92%	88.93%	99.95%	48
Molecular Index barcodes, +5%	97.22%	85.47%	99.93%	36
Molecular Index barcodes, +15%	100.00%	92.60%	100.00%	48
Molecular Index barcodes, +30%	100.00%	92.75%	100.00%	49

Taken together, these data indicate that the FoundationOne Liquid CDx assay is robust to potential specimen-related endogenous substances and exogenous contaminants or interferents.

### 9.7 Hybrid Capture Bait Specificity

Bait specificity was addressed through an assessment of coverage of targeted regions in FoundationOne Liquid CDx using 3,546 validation study samples. Results show that targeted genomic regions have consistently high, uniform coverage. For each genomic region associated with a predefined subset of highly-actionable alterations, between 94% to 100% of samples possessed the expected level of coverage. An in-depth, platform-wide examination of the FoundationOne Liquid CDx baitset through the analysis of HapMap process control samples revealed that, on average, 98.8% and 94.1% of platform-wide baited coding and non-coding regions, respectively, met their expected coverage levels. Samples assessed in this study consistently demonstrated high quality uniform and deep coverage across the entire genomic region targeted by the assay.

### 9.8 Carryover/Cross-Contamination

The study demonstrated that the risk of cross contamination (intra-plate), and carry-over contamination (inter-plate) of samples during the processing of the FoundationOne Liquid CDx assay is low. A total of 376 wells were examined for intra- and inter-plate contamination by processing and sequencing of contrived samples derived from cell lines at high input concentrations with known genomic backgrounds. Unique variants of each cell line were characterized by independent control sequencing runs. The samples were arrayed in a checkerboard fashion across four 96-well PCR plates to detect cross-contamination events. A cross-contamination rate of 0.53% (2/376) was observed in this study. These data demonstrate a low probability of cross contamination during the FoundationOne Liquid CDx process.

## 9.9 Precision: Repeatability and Reproducibility

Precision was evaluated for alterations associated with CDx claims, as well as tumor mutation profiling variants. Repeatability including intra-run performance (run on the same plate under the same conditions) and reproducibility including inter-run performance (run on different plates under different conditions) were assessed and compared across three reagent lots, two sequencers, and two processing runs.

### Results for a subset of highly-actionable alterations

A set of 39 unique samples were used to evaluate the precision of FoundationOne Liquid CDx for detecting a set of highly-actionable variants, including 8 contrived samples representing various targeted alterations and 31 clinical samples. The samples representing CDx alterations are summarized in **Table 19**. Additional non-CDx variants were evaluated as summarized in **Table 20**.

**Table 19: CDx sample set assessed for precision**

CDx Biomarker	Targeted Alteration	Disease Ontology of Patient from which Sample was Derived
ALK rearrangements	ALK-EML4 Rearrangement	Contrived sample
	ALK-EML4 Rearrangement	Lung adenocarcinoma
	ALK-NPM1 Rearrangement	Contrived sample
ATM alterations	ATM 5318delA	Contrived sample
	ATM I2012fs*4	Prostate cancer
	ATM splice site 8850+1G>A	Prostate cancer
	ATM-EXPH5 Truncation	Prostate cancer
BRCA1 and BRCA2 alterations	BRCA1 E23fs*17	Ovary cancer
	BRCA1 Q780*	Ovary high grade serous carcinoma
	BRCA1 Rearrangement	Unknown primary malignant neoplasm
	BRCA1_2475delC	Contrived sample
	BRCA1_2612C>TT	Contrived sample
	BRCA2_3599_3600delGT	Contrived sample
	BRCA2_4284_4285insT	Contrived sample
	BRCA2_5351delA	Contrived sample
	BRCA2 G267*	Ovary serous carcinoma
	BRCA2 Loss (15 of 26)	Prostate acinar adenocarcinoma
	BRCA2 Loss (26 of 26)	Prostate acinar adenocarcinoma
	BRCA2 S2988fs*12	Ovary cancer
BRCA2- EDA Truncation	Prostate cancer	
EGFR exon 19 deletions and EGFR exon 21 L858R alterations	EGFR E746_A750del	Non-small cell lung carcinoma
	EGFR_E746_A750del	Contrived sample
	EGFR L858R	Contrived sample
	EGFR L858R	Non-small cell lung carcinoma (2)
PIK3CA alterations	PIK3CA E542K	Contrived sample
	PIK3CA E542K, D549N	Contrived sample
	PIK3CA H1047R	Contrived sample
	PIK3CA E542K	Breast carcinoma
	PIK3CA E545K	Breast carcinoma
	PIK3CA H1047R	Breast cancer

**Table 20: Non-CDx sample set assessed for precision**

Non-CDx Targeted Alteration	Targeted Alteration	Disease Ontology of Patient from which Sample was Derived
<i>BRAF</i> alterations	<i>BRAF</i> L597R	Contrived sample
	<i>BRAF</i> V600E	Contrived sample
	<i>BRAF</i> V600E	Skin melanoma
	<i>BRAF</i> V600K	Skin melanoma
<i>EGFR</i> exon 20 T790M substitution	<i>EGFR</i> exon 20 T790M substitution	Contrived sample
<i>KRAS</i> alterations	<i>KRAS</i> G12D	Contrived sample
	<i>KRAS</i> G13D	Contrived sample
	<i>KRAS</i> G12L	Colon adenocarcinoma
	<i>KRAS</i> Q61R	Colon adenocarcinoma
<i>MET</i> exon 14 alterations	<i>MET</i> 3029-1G>T	Contrived sample
	<i>MET</i> 3933delC	Contrived sample
	<i>MET</i> exon 14 splice site 2888-17 2888-3del15	Non-small cell lung carcinoma
	<i>MET</i> exon 14 splice site 3005 3028+3>C	Non-small cell lung carcinoma
<i>NRAS</i> alterations	<i>NRAS</i> exon 2,3,4 substitutions	Contrived sample
<i>PALB2</i> alterations	<i>PALB2</i> 2422G>T	Contrived sample
	<i>PALB2</i> 2724delA	Contrived sample
<i>ERBB2</i> CNA	<i>ERBB2</i> CNA	Contrived sample
	<i>ERBB2</i> CNA	Breast carcinoma

Target alterations were assessed at two target levels each (near LoD and 2-3x LoD) for the contrived samples, and at one target level (1-1.5x LoD) for clinical cfDNA samples. Each sample was divided into 24 aliquots, with 12 duplicates being processed on the same plate under the same conditions. Across 47 samples (31 clinical specimens at one dilution level and 8 contrived samples across two dilution levels), a total of 57 unique alterations were evaluated. The repeatability and reproducibility of CDx alterations tested at >1x LoD is summarized in **Table 21**.

**Table 21 Repeatability and Reproducibility of CDx alterations targeted in precision study at >1x LoD\***

Variant Type	Alteration	Repeatability [%] {95% CI [%]}	Reproducibility [%] {95% CI [%]}	Level Tested**
<i>ATM</i> Short variant	<i>ATM</i> _5318delA	100 (73.54, 100)	100 (85.75, 100)	0.77% VAF
	<i>ATM</i> _5318delA	100 (71.51, 100)	100 (85.18, 100)	1.04% VAF
	<i>ATM</i> _6034_6035insCAGA AGTA	100 (71.51, 100)	100 (85.18, 100)	0.86% VAF
	<i>ATM</i> _8850+1G>A	100 (73.54, 100)	100 (85.75, 100)	0.56% VAF
<i>ATM</i> Rearrangement	<i>ATM-EXPH5</i> Truncation	100 (73.54, 100)	100 (85.75, 100)	1.13% VAF
<i>BRCA1</i> Short variant	<i>BRCA1</i> _2338C>T	100 (73.54, 100)	100 (85.75, 100)	1.11% VAF
	<i>BRCA1</i> _2475delC	100 (73.54, 100)	100 (85.75, 100)	0.61% VAF
	<i>BRCA1</i> _2475delC	100 (73.54, 100)	100 (85.75, 100)	0.93% VAF
	<i>BRCA1</i> _2612C>TT	100 (71.51, 100)	100 (85.18, 100)	0.51% VAF
	<i>BRCA1</i> _68_69delAG	100 (73.54, 100)	100 (85.75, 100)	0.66% VAF
	<i>BRCA1</i> _P871fs*32	100 (73.54, 100)	100 (85.75, 100)	1.08% VAF
<i>BRCA1</i> Rearrangement	<i>BRCA1-BRCA1</i>	100 (73.54, 100)	100 (85.75, 100)	0.87% VAF
<i>BRCA2</i> Short Variant	<i>BRCA2</i> _3599_3600delGT	100 (73.54, 100)	100 (85.75, 100)	0.58% VAF
	<i>BRCA2</i> _3599_3600delGT	100 (73.54, 100)	100 (85.75, 100)	0.92% VAF

Variant Type	Alteration	Repeatability [%] {95% CI [%]}	Reproducibility [%] {95% CI [%]}	Level Tested**
	BRCA2_4284_4285insT	100 (73.54, 100)	100 (85.75, 100)	0.94% VAF
	BRCA2_4284_4285insT	100 (71.51, 100)	100 (85.18, 100)	1.26% VAF
	BRCA2_5351delA	100 (73.54, 100)	100 (85.75, 100)	1.22% VAF
	BRCA2_5351delA	100 (73.54, 100)	100 (85.75, 100)	1.85% VAF
	BRCA2_5351delA	100 (71.51, 100)	100 (85.18, 100)	1.07% VAF
	BRCA2_5351delA	100 (73.54, 100)	100 (85.75, 100)	2.24% VAF
	BRCA2_5465_5466insA	100 (73.54, 100)	100 (85.75, 100)	0.92% VAF
	BRCA2_5465_5466insA	100 (71.51, 100)	100 (85.18, 100)	1.19% VAF
	BRCA2_8961_8964delGA GT	100 (73.54, 100)	100 (85.75, 100)	1.07% VAF
	BRCA2_c.799G>T	83.33 (51.59, 97.91)	91.67 (73.0, 98.97)	0.5% VAF
	BRCA2_c.9097_9098insA	54.55 (23.38, 83.25)	21.74 (7.46, 43.7)	0.71% VAF
	BRCA2_c.9097_9098insA	83.33 (51.59, 97.91)	91.67 (73.0, 98.97)	1.03% VAF
	BRCA2 Copy Number Loss	BRCA2_loss	91.67 (61.52, 99.79)	87.5 (67.64, 97.34)
BRCA2 Rearrangement	BRCA2-EDA	100 (71.51, 100)	100 (85.18, 100)	0.48% VAF
EGFR Short variant	EGFR_2369C>T	100 (73.54, 100)	100 (85.75, 100)	0.44% VAF
	EGFR_2369C>T	100 (73.54, 100)	100 (85.75, 100)	0.66% VAF
	EGFR_2369C>T	100 (71.51, 100)	100 (85.18, 100)	0.36% VAF
	EGFR_2369C>T	100 (73.54, 100)	100 (85.75, 100)	0.65% VAF
	EGFR_2369C>T	100 (73.54, 100)	100 (85.75, 100)	1.26% VAF
	EGFR_2573T>G	100 (73.54, 100)	100 (85.75, 100)	0.46% VAF
	EGFR_2573T>G	100 (73.54, 100)	100 (85.75, 100)	0.68% VAF
	EGFR_2573T>G	100 (73.54, 100)	100 (85.75, 100)	0.68% VAF
	EGFR_2573T>G	100 (71.51, 100)	100 (85.18, 100)	0.95% VAF
	EGFR_2573T>G	100 (73.54, 100)	100 (85.75, 100)	0.64% VAF
	EGFR_2573T>G	100 (73.54, 100)	100 (85.75, 100)	1.64% VAF
	EGFR_E746_A750del	100 (73.54, 100)	100 (85.75, 100)	0.51% VAF
	EGFR_E746_A750del	100 (73.54, 100)	100 (85.75, 100)	0.74% VAF
	EGFR_E746_A750del	100 (73.54, 100)	100 (85.75, 100)	0.93% VAF
	EGFR_E746_A750del	100 (71.51, 100)	100 (85.18, 100)	1.2% VAF
	EGFR_E746_A750del	100 (71.51, 100)	100 (85.18, 100)	0.51% VAF
	EGFR_E746_A750del	100 (73.54, 100)	100 (85.75, 100)	1.01% VAF
EGFR_E746_A750del	100 (71.51, 100)	100 (84.56, 100)	0.34% VAF	
PIK3CA Short variant	PIK3CA_1624G>A	100 (73.54, 100)	100 (85.75, 100)	0.89% VAF
	PIK3CA_1633G>A	100 (73.54, 100)	100 (85.75, 100)	0.45% VAF
	PIK3CA_1633G>A	100 (73.54, 100)	100 (85.75, 100)	0.66% VAF
	PIK3CA_1633G>A	100 (73.54, 100)	100 (85.75, 100)	0.5% VAF
	PIK3CA_1634A>C	100 (73.54, 100)	100 (85.75, 100)	0.52% VAF
	PIK3CA_1634A>C	100 (71.51, 100)	100 (85.18, 100)	0.70% VAF
	PIK3CA_1637A>G	90.91 (58.72, 99.77)	95.65 (78.05, 9.89)	0.49% VAF
	PIK3CA_1637A>G	100 (73.54, 100)	100 (85.75, 100)	0.92% VAF
	PIK3CA_1645G>A	100 (73.54, 100)	100 (85.75, 100)	0.48% VAF
	PIK3CA_1645G>A	100 (73.54, 100)	100 (85.75, 100)	0.73% VAF
	PIK3CA_3140A>G	100 (71.51, 100)	100 (85.18, 100)	0.41% VAF
	PIK3CA_3140A>G	100 (73.54, 100)	100 (85.75, 100)	0.76% VAF
	PIK3CA_3140A>G	100 (73.54, 100)	100 (85.75, 100)	1.04% VAF
ALK Rearrangement	ALK_EML4	100 (73.54, 100)	100 (85.75, 100)	0.64% VAF
	ALK_EML4	100 (71.51, 100)	100 (85.18, 100)	0.89% VAF
	ALK_EML4	100 (73.54, 100)	100 (85.75, 100)	1.39% VAF
	ALK-NPM1	100 (73.54, 100)	100 (85.75, 100)	0.64% VAF

\*Clinical samples were mostly tested at 2x – 3x LoD rather than 1x – 1.5x LoD

\*\*The accuracy of %VAF/%TF have not been analytically validated

As observed in the **Table 21** above, three BRCA2 positive samples (c.799G>T, c.9097\_9098insA, and a BRCA2 loss) demonstrated poor performance for both repeatability and reproducibility. For the BRCA2

specimen harboring the c.799G>T, the average %VAF was determined to be 0.5%, near the LoD of 0.4% for this variant type. The *BRCA2* c.9097\_9098insA variant is an insertion of an A in a highly repetitive homopolymer region of eight As, which impacts sensitivity. In the LoD study, a 93% hit rate was observed at the highest level tested, 1.16% VAF, indicating that the levels evaluated in this precision analysis were below the LoD for this variant. The replicates for the clinical sample harboring the *BRCA2* loss were processed at below the minimum cfDNA input.

Of 53 targeted alterations, repeatability of 100% was observed for 43 alterations and ≥90% repeatability was observed for 53 alterations. For the targeted variants assessed, the overall repeatability was 96.39% (95% two-sided exact CIs [95.28%, 97.30%]).

Of 55 targeted alterations, reproducibility of 100% was observed for 42 alterations and ≥90% reproducibility was observed for 55 alterations. For the targeted variants assessed, the overall reproducibility was 97.33% (95% 2-sided exact CIs [96.67 %, 97.89%]).

The repeatability and reproducibility of non-CDx alterations tested at ≥1x LoD are summarized in **Table 22**.

**Table 22: Repeatability and Reproducibility of non-CDx alterations targeted in precision study at ≥1x LoD**

Variant Type	Alteration	Repeatability [%] {95% CI [%]}	Reproducibility [%] {95% CI [%]}	Level Tested*
<i>BRAF</i> Short variant	BRAF_1790T>G	90.91 (58.72, 99.77)	95.65 (78.88, 99.89)	0.42% VAF
	BRAF_1790T>G	100 (73.54, 100)	100 (85.75, 100)	0.85% VAF
	BRAF_1798_1799GT>AA	91.67 (61.52, 99.79)	95.83 (78.88, 99.89)	0.36% VAF
	BRAF_1799T>A	100 (71.51, 100)	100 (85.18, 100)	0.72% VAF
	BRAF_1799T>A	100 (73.54, 100)	100 (85.75, 100)	1.38% VAF
	BRAF_1799T>A	100 (73.54, 100)	100 (85.75, 100)	0.44% VAF
<i>KRAS</i> Short variant	KRAS_182A>G	100 (73.54, 100)	100 (85.75, 100)	0.53% VAF
	KRAS_34_35GG>CT	100 (73.54, 100)	100 (85.75, 100)	0.49% VAF
	KRAS_35G>A	100 (73.54, 100)	100 (85.75, 100)	0.89% VAF
	KRAS_35G>A	100 (71.51, 100)	100 (85.18, 100)	1.12% VAF
	KRAS_38G>A	100 (73.54, 100)	100 (85.75, 100)	0.55% VAF
	KRAS_38G>A	100 (73.54, 100)	100 (85.75, 100)	0.82% VAF
	KRAS_38G>A	100 (71.51, 100)	100 (85.18, 100)	0.57% VAF
	KRAS_38G>A	100 (73.54, 100)	100 (85.75, 100)	0.92% VAF
<i>MET</i> Short variant	MET_2888-17_2888-3del15	100 (73.54, 100)	100 (85.75, 100)	1.17% VAF
	MET_3005_3028+3>C	100 (73.54, 100)	100 (85.75, 100)	1.67% VAF
	MET_3029-1G>T	81.82 (48.22, 97.72)	91.30 (71.96, 98.93)	0.30% VAF
	MET_3933delC	100 (73.54, 100)	100 (85.75, 100)	0.69% VAF
	MET_3933delC	100 (73.54, 100)	100 (85.75, 100)	0.96% VAF
<i>NRAS</i> Short variant	NRAS_34G>T	100 (73.54, 100)	100 (85.75, 100)	0.69% VAF
	NRAS_34G>T	100 (73.54, 100)	100 (85.75, 100)	0.96% VAF
	NRAS_35G>A	100 (73.54, 100)	100 (85.75, 100)	0.84% VAF
	NRAS_c.35G>A	63.64 (30.79, 89.07)	82.61 (61.22, 95.05)	0.48% VAF
<i>PALB2</i> Short variant	PALB2_2422G>T	100 (71.51, 100)	100 (85.18, 100)	0.47% VAF
	PALB2_2422G>T	100 (73.54, 100)	100 (85.75, 100)	0.92% VAF
	PALB2_2724delA	100 (73.54, 100)	100 (85.75, 100)	0.52% VAF

Variant Type	Alteration	Repeatability [%] {95% CI [%]}	Reproducibility [%] {95% CI [%]}	Level Tested*
	PALB2_2724delA	100 (73.54, 100)	100 (85.75, 100)	0.74% VAF
ERBB2 CN Amplification	ERBB2 amplification	100 (73.54, 100)	100 (85.75, 100)	35.78% VAF
	ERBB2 amplification	100 (73.54, 100)	100 (85.75, 100)	39.79% VAF
	ERBB2 amplification	100 (73.54, 100)	100 (85.75, 100)	61.73% VAF

\*The accuracy of %VAF/%TF have not been analytically validated

### Precision of Platform Variants

Across 39 unique samples, including 8 contrived samples, and 31 clinical samples, a total of 1,240 variants were evaluated with variant types including substitutions, indels, rearrangements, and copy number alterations. The number of variants in each variant bin are summarized in **Table 23**.

**Table 23: Number of each variant type**

Variant Category	N
<b>Substitutions</b>	<b>898</b>
Substitution in a non-repetitive region or a repetitive region of <=7 base pairs	882
Substitution in a repetitive region of >7 base pairs	16
<b>Indels</b>	<b>228</b>
Insertion/Deletion in non-repetitive region or a repetitive region of <=3 base pairs	52
Insertion/Deletion in a repetitive region of 4 to 6 base pairs	118
Insertion/Deletion in a repetitive region of >=7 base pairs	58
<b>Rearrangements</b>	<b>60</b>
<b>Copy Number Alterations</b>	<b>54</b>
Copy Number Amplification	49
Copy Number Loss	5
<b>Total</b>	<b>1240</b>

The overall repeatability for all variants were 99.47% with 95% 2-sided exact CIs (99.45%, 99.48%). The repeatability result for each variant type are summarized in **Table 24**.

**Table 24: Assessment of repeatability of tumor mutation profiling variants per type**

Variant Type	# of Concordant Pairs	# of Total Pairs	Repeatability (%)	95% two-sided exact CIs (%)
Substitution	498765	501084	99.54	(99.52, 99.56)
Indels	126475	127224	99.41	(99.37, 99.45)
Rearrangements	33105	33480	98.88	(98.76, 98.99)
Copy Number Alterations	29880	30132	99.16	(99.05, 99.26)

The overall reproducibility results were 99.59% with the 95% 2-sided exact CIs (99.58%, 99.60%). The reproducibility result for each variant type are summarized in **Table 25**.

**Table 25: Assessment of reproducibility of tumor mutation profiling variants per type**

Variant Type	# of Concordant Replicates	# of Total Replicates	Reproducibility (%)	95% two-sided exact CIs (%)
Substitution	1002981	1006658	99.63	(99.62, 99.65)
Indels	254509	255588	99.58	(99.55, 99.60)
Rearrangements	66723	67260	99.20	(99.13, 99.27)
Copy Number Alterations	60115	60534	99.31	(99.24, 99.7)

**Confirmation of LoD and Precision in Clinical Specimens**

Twenty-nine clinical cfDNA samples targeting variants at near the estimated LoD were evaluated to confirm LoD and precision in clinical specimens. The mean level tested in most cases were higher than the estimated LoD as shown in **Table 26** and **Table 27**. Twenty-six had 100% reproducibility, one had 95.8% reproducibility, and two samples had reproducibility below 90%. Of these two samples, one contained a *BRCA2* loss that had 87.5% reproducibility. This sample was processed with a cfDNA input mass below the recommended minimum and was also below LoD. The other sample harbored a *BRCA2* substitution (c.799G>T) with 91.67% reproducibility. The average VAF of this variant was 0.5% across replicates, which is near the LoD for this variant type (median LoD of 0.4% VAF). A summary of the Confirmation of LoD and Precision results for CDx variants are provided in **Table 26**. A summary of the Confirmation of LoD and Precision results for CDx variants are provided in **Table 27**.

**Table 26: CDx variant confirmation of LoD and precision in clinical specimens**

Target Alteration	LoD	Mean Level Tested*	Reproducibility (%)	95% Two-sided exact CIs (%)
<i>ATM</i> I2012fs*4	0.51% VAF	0.86% VAF	100	(85.18, 100)
<i>ATM</i> splice site 8850+1G>A	0.51% VAF	0.56% VAF	100	(85.75, 100)
<i>BRCA1</i> E23fs*17	0.38% VAF	0.66% VAF	100	(85.75, 100)
<i>BRCA1</i> Q780*	0.34% VAF	1.11% VAF	100	(85.75, 100)
<i>BRCA1</i> Rearrangement	0.87% VAF <sup>1</sup>	0.87% VAF	100	(85.75, 100)
<i>BRCA2</i> 799G>T	0.40% VAF	0.50% VAF	91.67	(73.0, 98.97)
<i>BRCA2</i> Loss	48.1% TF	39.43% TF	87.50	(67.64, 97.34)
<i>BRCA2</i> S2988fs*12	0.36% VAF	1.07% VAF	100	(85.75, 100)
<i>BRCA2</i> - EDA Truncation	0.48% VAF <sup>1</sup>	0.48% VAF	100	(85.18, 100)
<i>EGFR</i> E746_A750del	0.27% VAF	0.34% VAF	100	(84.56, 100)
<i>EGFR</i> L858R	0.34% VAF	1.64% VAF	100	(85.75, 100)
<i>EGFR</i> L858R	0.34% VAF	0.64% VAF	100	(85.75, 100)
<i>PIK3CA</i> E542K	0.34% VAF	0.89% VAF	100	(85.75, 100)
<i>PIK3CA</i> E545K	0.34% VAF	0.5% VAF	100	(85.75, 100)
<i>PIK3CA</i> H1047R	0.34% VAF	1.04% VAF	100	(85.75, 100)
<i>ALK-EML4</i> Rearrangement	0.24% MAF	1.39% MAF	100	(85.75, 100)

<sup>1</sup> LoD determined in this confirmation of LoD and precision study

\*The accuracy of %VAF/%TF have not been analytically validated



**Table 27: Non-CDx variant confirmation of LoD and precision in clinical specimens**

Target Alteration	LoD	Mean Level Tested*	Reproducibility (%)	95% Two-sided exact CIs (%)
<i>BRAF</i> V600E	0.33% VAF	0.44% VAF	100	(85.75, 100)
<i>BRAF</i> V600K	0.33% VAF	0.36% VAF	95.8	(78.88, 99.89)
<i>EGFR</i> T790M	0.34% VAF	1.26% VAF	100	(85.75, 100)
<i>KRAS</i> G12L	0.33% VAF	0.49% VAF	100	(85.75, 100)
<i>KRAS</i> Q61R	0.33% VAF	0.53% VAF	100	(85.75, 100)
<i>MET</i> exon 14 splice site 2888-17_2888-3del15	0.41% VAF	1.17%	100	(85.75, 100)
<i>MET</i> exon 14splice site 3005_3028+3>C	0.41% VAF	1.67% VAF	100	(85.75, 100)
<i>ERBB2</i> CNA	19.8% TF	61.73% TF	100	(85.75, 100)

\*The accuracy of %VAF/%TF have not been analytically validated

A second study with 10 samples targeting variants at 1-1.5x LoD was performed to confirm LoD and precision in clinical specimens. Similar to above, each sample was divided into 24 aliquots, with 12 duplicates being processed on the same plate under the same conditions. Each sample was tested across 24 replicates. Six samples were included in the primary analysis for samples with  $\geq 30$  ng DNA input. Three had 100% reproducibility, one had 95.7% reproducibility, one had 91.7% reproducibility, and one had 91.3% reproducibility. The other four samples had a majority of sample replicates with DNA input  $< 30$  ng. A summary of the Confirmation of LoD and Precision results for CDx alterations are provided in **Table 28**.

**Table 28: CDx variant confirmation of LoD and precision in clinical specimens**

Target Alteration	LoD	Mean Level Tested*	Reproducibility (95% CI)	95% CIs (%)
<i>BRCA1</i> 1395T>A	0.34%	0.51%	100%	[86.2%, 100%]
<i>BRCA2</i> 5351_5352insA	0.36%	0.34%	87.5%	[69.0%, 95.7%]
<i>EGFR</i> 2235_2249del	0.27%	0.45%	95.7%	[79.0%, 99.2%]
<i>PIK3CA</i> 1637A>G	0.34%	0.44%	91.7%	[74.2%, 97.7%]

\*The accuracy of %VAF/%TF have not been analytically validated

As summarized in **Table 28** above, all CDx variants with  $\geq 30$  ng DNA input had reproducibility  $\geq 95\%$  with the exception of one variant (*BRCA2* 5351\_5352insA) which was tested at a variant allele fraction below the LoD.

Additionally, one of the 10 samples evaluated in this study targeted a non-CDx *BRCA2* substitution. Reproducibility of 100% was observed as summarized in **Table 29**.

**Table 29: Non-CDx variant confirmation of LoD and precision in a clinical specimen**

Target Alteration	LoD	Mean Level Tested*	Reproducibility (95% CI)	95% CIs (%)
<i>BRCA2</i> 8524C>T	0.37%	0.57%	100%	[85.7%, 100%]
<i>NRAS</i> 34G>T	0.42%	0.55%	91.3%	[73.2%, 97.6%]

\*The accuracy of %VAF/%TF have not been analytically validated

## 9.10 Reagent Lot Interchangeability

The interchangeability of critical reagent lots for library construction (LC), hybrid capture (HC) and sequencing within the FoundationOne Liquid CDx assay was evaluated by testing eight (8) contrived samples from either

enzymatically fragmented cell line genomic DNA containing alterations of interest or enzymatically fragmented plasmid DNA. Each of the contrived samples was tested in triplicate using two different lots each of LC, HC, and sequencing reagents. Eight reagent pairings were assessed. A total of eight analyses for each specimen were completed. 192 tests in total were included in this study. Four Master Pool Libraries (MPLs) were evaluated on each of two flowcells on a NovaSeq 6000 sequencer, using two different Sequencing reagent lots. Of the 49 alterations assessed in the sample set, 43 had a percent agreement greater than 90% (39 alterations had percentage agreement equal to 100%, one had percent agreement equal to 95.83%, one had percent agreement equal to 95.65%, and two had percent agreement equal to 91.67%), exceeding the pre-specified acceptance criteria. For the remaining six alterations the observed detection rates for these variants were similar to the predicted detection rate based on the LoD analysis. These results demonstrate the interchangeability of critical reagent lots in the FoundationOne Liquid CDx assay.

### 9.11 Variant Curator Precision

This study was performed to evaluate the precision of genomic variant call curation, following analysis by the FoundationOne Liquid CDx analysis pipeline. This was established by analyzing targeted alterations, including CDx alterations, and platform-wide alterations within samples used in the FoundationOne Liquid CDx Precision and LoD and Precision Confirmation Study. The study design reflected the intermediate precision design and evaluated curator precision in reporting of targeted and platform alterations. A total of 19 samples were selected for this study. Three curators were chosen randomly amongst all qualified curators to curate variant calls in a set of randomly chosen replicates from each of the 19 samples. The variant calls were generated from each sample per curator. The overall average percent agreement for targeted alterations was 93.3% (95% CI; 83.80%, 98.15%), and for platform genomic alterations was 99.14% (95% CI; 98.47%, 99.57%).

### 9.12 Stability

#### 9.12.1 Reagent Stability

The reagent stability of FoundationOne Liquid CDx is assessed by analyzing data from each of eight samples in triplicate, per each of three different lots of LC, HC, and sequencing reagents. A total of nine analyses for each specimen are completed for each of six time points assessed. A total of 72 tests will be assessed per time period; a total of 432 samples and six time points will be included in this study overall. Each of the three sample Master Library Pools (MPLs), representing three LC and HC reagent lots will be evaluated per time point on a NovaSeq 6000 sequencer, using three different sequencing reagent lots. The analysis of baseline timepoint zero (T0) identified the baseline variant calls for each sample. Concordance of 12,511 variant alterations will be assessed across future time points for sample aliquots derived from eight DNA samples.

To date, timepoint the 3-month timepoint has been analyzed for reagent Lot #1, Lot #2, and Lot #3. Variants at the experimental time points are  $\geq 90\%$  concordant with the baseline variant call values as presented in **Table 30**. Current data demonstrate LC, HC, and sequencing reagent stability for up to 3 months. This study is ongoing and further evaluation will be performed to validate reagent stability over 12 months.

**Table 30: Concordance analysis between 3 months and baseline**

	Reagent Lot	Timepoint		Total # Replicates	Concordance Percentage	95% C.I.	
Variant Calls	Lot #1	1	1921	1966	97.71%	96.95%	98.28%
	Lot #2	1	2083	2148	96.97%	96.16%	97.62%
	Lot #3	1	2086	2139	97.52%	96.77%	98.10%

#### 9.12.2 Whole Blood Specimen Stability

Whole blood stability and the impact of tube inversion was evaluated in freshly collected whole blood samples from the following five cancer types: non-small cell lung cancer (NSCLC), colorectal cancer (CRC), prostate, breast, and ovarian cancer. The recommended storage temperature is 18°C - 25°C. In this study, stress conditions were simulated through extended storage at elevated (35°C  $\pm$  2°C) and reduced (4°  $\pm$  2°C) temperatures.

In this interim analysis, 22 samples (11 sample pairs) were tested, including baseline (within 24 hours of collection) and experimental time points (after 10, 14, or 15 days of storage).

Overall, 100% of samples yielded a cfDNA input  $\geq 30$ ng. The success rate for DNAX yield, and LC yield were 100% and the success rate of the HC yield was 96.3%. The variant analysis was conducted for variants at  $\geq 2$ x LoD. For the aggregate 11 pairs of samples processed and reported, 100% agreement was observed between the baseline and experimental timepoint for short variants and rearrangements for each experimental time point. The percent agreement per sample also resulted in 100% agreement between the baseline and experimental timepoint for short variants and rearrangements. The data is summarized in **Table 31**.

**Table 31: Aggregate percent agreement per temperature and experimental timepoint**

Temperature	Experimental Timepoint	N	Short Variants [95% two-sided CI]	Rearrangements
4°C	7 Days	4	100.00 [89.72, 100.00]	100.00 [39.76, 100.00]
	14 Days	3	100.00 [91.40, 100.00]	N/A
	15 Days	3	100.00 [83.89, 100.00]	N/A
35°C	14 Days	1	N/A	N/A

The impact of potential interferents originating from the FoundationOne Liquid cfDNA blood collection tube (BCT) stopper on the performance of the FoundationOne Liquid CDx assay was assessed by evaluating stability of whole blood in tubes stored in an upright or inverted position at  $4^{\circ}\text{C} \pm 2^{\circ}\text{C}$ ,  $25^{\circ}\text{C} \pm 2^{\circ}\text{C}$ , and  $35^{\circ}\text{C} \pm 2^{\circ}\text{C}$  for various durations (10, 14, and 15 days).

First, the success rate of the FoundationOne Liquid CDx assay for processing samples was assessed at the DNA extraction (DNAX), Library Construction (LC), Hybrid Capture (HC) and Sequencing step, based on product in-process quality control (QC) criteria. Samples stratified by the upright and the inverted condition exhibited comparable success rates above 94% at DNAX, LC, HC and Seq (**Table 32**). Thus, the stopper of the FoundationOne Liquid cfDNA BCT does not impact FoundationOne Liquid CDx test performance when stored between 4 and  $35^{\circ}\text{C}$  for up to 15 days.

**Table 32: Process success rate by tube position**

Process	Tube Position	# Passing Samples	# Total Samples	Success Rate (%)	95% 2-sided CIs (%)
DNA Extraction	Upright	139	147	94.6%	[89.6%, 97.2%]
	Inverted	147	150	98%	[94.3%, 99.3%]
LC	Upright	135	136	99.3%	[96%, 99.9%]
	Inverted	146	146	100%	[97.4%, 100%]
HC	Upright	134	135	99.3%	[95.9%, 99.9%]
	Inverted	143	146	97.9%	[94.1%, 99.3%]
Sequencing	Upright	134	134	100%	[97.2%, 100%]
	Inverted	143	143	100%	[97.4%, 100%]

Stability was also evaluated by comparing concordance between baseline and experimental samples. Positive percent agreement (PPA) and negative percent agreement (NPA) for alteration calls at  $\geq 2$ x LoD were computed along with the corresponding two-sided 95% score confidence interval (CI) across all replicates by variant

category using the baseline detection as reference. Note that NPA is under-estimated as variants not detected at any of the treatment conditions were not used in the analysis set and hence counted against the NPA calculation.

Concordance between baseline and experimental results from all samples in the upright and inverted position combined demonstrated > 99% PPA and NPA for the detection of short variants and rearrangements. Copy number alterations were only detected in samples treated in the inverted tube position and therefore, not included in this analysis. Furthermore, stratification by the treatment condition (2 tube positions × 3 temperatures × 3 durations) revealed >99.0% PPA and NPA for short variants and rearrangements across the combinations of tube positions, temperatures and durations tested. The data also demonstrate that the detection of copy number alterations is not impacted by the storage of blood in the inverted position at 35°C for up to 14 days. The concordance results by variant type for each of the experimental conditions are provided in **Table 33**.

**Table 33: Concordance of detected alterations between baseline sample and experimental conditions for inverted tube stability study**

Variant Type	Temp.	Tube Position	Exp. Time Point	N Variants Detected at Baseline Time Point	N Variants Detected at Exp. Time Point	N Variants Agree	PPA	PPA [95% CI]	N Variants Not Detected at Baseline Time Point	N Variants Not Detected at Exp. Time Point	NPA	NPA [95% CI]
Short variants	04°C	Inverted	Day 10	50	50	49	98%	[89.5%, 99.6%]	612	612	100%	[100%, 100%]
Short variants	04°C	Upright	Day 10	50	51	50	100%	[92.9%, 100%]	613	612	100%	[100%, 100%]
Short variants	04°C	Inverted	Day 14	59	58	58	98.3%	[90.9%, 99.7%]	610	611	100%	[100%, 100%]
Short variants	04°C	Upright	Day 14	44	44	44	100%	[92.0%, 100%]	611	611	100%	[100%, 100%]
Short variants	04°C	Inverted	Day 15	37	37	37	100%	[90.6%, 100%]	611	611	100%	[100%, 100%]
Short variants	04°C	Upright	Day 15	52	52	52	100%	[93%, 100%]	611	611	100%	[100%, 100%]
Short variants	25°C	Inverted	Day 10	78	77	76	97.1%	[91.1%, 99.2%]	627	628	100%	[100%, 100%]
Short variants	25°C	Upright	Day 10	44	44	44	100%	[92.0%, 100%]	613	613	100%	[100%, 100%]
Short variants	25°C	Inverted	Day 14	46	48	46	100%	[92.3%, 100%]	611	609	100%	[100%, 100%]
Short variants	25°C	Upright	Day 14	42	41	41	97.6%	[87.7%, 99.6%]	610	611	100%	[100%, 100%]
Short variants	25°C	Inverted	Day 15	44	44	44	100%	[92.0%, 100%]	613	613	100%	[100%, 100%]
Short variants	25°C	Upright	Day 15	49	48	48	97.8%	[89.3%, 99.6%]	616	617	100%	[100%, 100%]
Short variants	35°C	Inverted	Day 10	15	15	15	100%	[79.6%, 100%]	609	609	100%	[100%, 100%]
Short variants	35°C	Upright	Day 10	35	35	35	100%	[90.1%, 100%]	609	609	100%	[100%, 100%]
Short variants	35°C	Inverted	Day 14	55	55	55	100%	[93.4%, 100%]	611	611	100%	[100%, 100%]
Short variants	35°C	Upright	Day 14	48	47	46	95.7%	[86.0%, 98.8%]	609	610	100%	[100%, 100%]

Variant Type	Temp.	Tube Position	Exp. Time Point	N Variants Detected at Baseline Time Point	N Variants Detected at Exp. Time Point	N Variants Agree	PPA	PPA [95% CI]	N Variants Not Detected at Baseline Time Point	N Variants Not Detected at Exp. Time Point	NPA	NPA [95% CI]
Short variants	35°C	Inverted	Day 15	39	39	38	97.4%	[86.8%, 99.5%]	610	610	100%	[100%, 100%]
Short variants	35°C	Upright	Day 15	28	29	28	100%	[87.9%, 100%]	613	612	100%	[100%, 100%]

These results demonstrate that blood is stable in the FoundationOne Liquid CDx cfDNA BCT when stored between 4°C and 35°C for up to 15 days, in an upright or inverted position. Additional data will be generated to further evaluate whole blood stability and potential interference of the blood collection tube cap.

### 9.13 DNA Extraction

DNA extraction evaluated 72 samples across five cancer types: lung cancer (including NSCLC), colorectal cancer (CRC), prostate cancer, breast cancer, and skin cancer (melanoma, sarcoma), using three reagent lots and two KingFisher Magnetic Particle processors.

Reproducibility of the FoundationOne Liquid CDx DNA extraction process across KingFisher instruments and extraction reagent lots were analyzed utilizing a factorial design (3 reagent lots × 2 KingFisher instruments × 2 replicates). The success rate of the DNAX yield for three reagent lots range from 95.8% to 100.0% and two King Fisher instruments range from 97.2% to 100.0%.

Variant calls included in the concordance analysis were identified based on the majority call across all 12 replicates for a given disease ontology. Positive Percent Agreement (PPA) and Negative Percent Agreement (NPA) were computed across the replicates for each somatic alteration for each sample, and aggregated by variant type (deletion, insertion, rearrangement, and substitution) for variants at ≥1x LoD. The percent agreement results by disease ontologies are: 90.3% - 99.8 % for PPA, and 99.1% - 100.0% for NPA (**Table 34**). The percent agreement results across all variant types (deletion, insertion, rearrangement and substitution) evaluated at ≥1x LoD are: 90.6% - 96.8% for PPA and 98.9% - 100.0% for NPA (**Table 35**).

**Table 34: Concordance summary by disease ontology at 1x LoD for DNA extraction study**

Disease Ontology	Positive Detected/ Positive Total	PPA [95% two-sided CI]	Negative Detected/ Negative Total*	NPA [95% two-sided CI]	Overall Detected/ Total*	OPA [95% two-sided CI]
Breast Cancer	347/348	99.7% [98.4%,100.0%]	3144/3144	100.0% [99.9%,100.0%]	3491/3492	100.0% [99.8%,100.0%]
Colorectal Cancer (CRC)	1122/1188	94.4% [93.0%,95.7%]	2284/2304	99.1% [98.7%,99.5%]	3406/3492	97.5% [97.0%,98.0%]
Lung Cancer	431/432	99.8% [98.7%,100.0%]	3053/3060	99.8% [99.5%,99.9%]	3484/3492	99.8% [99.5%,99.9%]
Non-Small Cell Lung Cancer (NSCLC)	600/612	98.0% [96.6%,99.0%]	2878/2880	99.9% [99.7%,100.0%]	3478/3492	99.6% [99.3%,99.8%]
Prostate Cancer	486/492	98.8% [97.4%,99.6%]	2987/3000	99.6% [99.3%,99.8%]	3473/3492	99.5% [99.2%,99.7%]

Skin Cancer	455/504	90.3% [87.4%,92.7%]	2987/2988	100.0% [99.8%,100.0%]	3442/3492	98.6% [98.1%,98.9%]
-------------	---------	------------------------	-----------	--------------------------	-----------	------------------------

\* Variants detected include variants classified as VUS and benign

**Table 35: Concordance summary by variant type at 1x LoD for DNA extraction study**

Variant Type	Positive Detected/ Positive Total	PPA [95% two-sided CI]	Negative Detected/ Negative Total*	NPA [95% two-sided CI]	Overall Detected/ Total*	OPA [95% two-sided CI]
<b>Deletions</b>	386/ 408	94.6% [91.9%, 96.6%]	2036/ 2040	99.8% [99.5% 99.9%]	2422/ 2448	98.9% [98.4% 99.3%]
<b>Insertions</b>	163/ 180	90.6% [85.3%, 94.4%]	819/ 828	98.9% [97.9% 99.5%]	982/ 1008	97.4% [96.2% 98.3%]
<b>Rearrangements</b>	23/ 24	95.8% [78.9%, 99.9%]	120/ 120	100.0% [97.0% 100.0%]	143/ 144	99.3% [96.2% 100.0%]
<b>Substitutions</b>	2869/ 2964	96.8% [96.1%, 97.4%]	14358/ 14388	99.8% [99.7% 99.9%]	17227/ 17352	99.3% [99.1% 99.4%]

\* Variants detected include variants classified as VUS and benign

These results demonstrate robustness of the FoundationOne Liquid CDx DNA extraction process across KingFisher instruments, extraction reagent lots, and cancer types.

#### 9.14 Guard Banding/Robustness

The purpose of this validation study was to evaluate the impact on FoundationOne Liquid CDx test performance due to potential process variation with regard to uncertainty in the measurement of DNA concentration. This guard banding evaluation assessed the DNA input into each of the main process steps of the FoundationOne Liquid CDx assay (LC, HC, and sequencing).

Guard bands were evaluated relative to calculated process variability for LC, HC, and sequencing. The assessment of multiple DNA input levels into LC demonstrated robust performance and tolerance of various DNA input levels. The observed results of HC guard banding showed that the HC process is robust within the predefined specifications 1000ng to 2000ng of DNA input into HC. For sequencing, the observed distribution of coverage indicated robust performance within the predefined specifications of 1.0nM of DNA input concentration into sequencing (as summarized in **Table 36**).

**Table 36: Summary of process pass and failure rate at each guard banding DNA input level**

Process	Input Level	# of Pass	Pass Rate (%)
LC	-33%	20ng	20/20
	-20%	24ng	20/20
	Recommended lower limit	30ng	20/20
	Low input	45ng	20/20
	Mid-point	55ng	20/20
	Upper limit	80ng	20/20
	+20%	96ng	19/20*
	+33%	106ng	20/20

HC	-50%	500ng	18/20	90
	-20%	800ng	20/20	100
	Lower limit	1000ng	20/20	100
	Upper limit	2000ng	20/20	100
	+20%	2400ng	20/20	100
	+50%	3000ng	18/20	90
Sequencing	-50%	0.5nM	20/20	100
	-20%	0.8nM	20/20	100
	Normal input	1.0nM	20/20	100
	+20%	1.2nM	20/20	100
	+50%	1.5nM	20/20	100

\* This one failure was due to failure of HC PICO DNA yield rather than LC PICO DNA yield.

### 9.15 Pan-Tumor Performance

A large-scale retrospective analysis was performed to demonstrate consistent test performance of FoundationOne Liquid CDx across samples derived from patients with different tumor types. This was evaluated by comparing in-process QC metrics across tumor types using historical data from samples processed in Foundation Medicine's clinical laboratory using two prior versions of the FoundationOne Liquid CDx assay. The FoundationOne Liquid CDx assay was developed based on two versions of the FoundationOne Liquid LDT assay, each of which targeted a subset of the genomic regions targeted by FoundationOne Liquid CDx. FoundationACT (FACT) targeted 62 genes and FoundationOne Liquid targeted 70 genes. The workflow is substantially similar between the assays. In order to support the use of historical data in this study, the regions commonly baited by the two previous assay versions and by FoundationOne Liquid CDx were evaluated for comparability of test performance (Section 2.15).

The sample set for this analysis included 19,868 distinct samples from 25 tumor type categories that had previously been tested using the Foundation Medicine FoundationOne Liquid and FoundationACT assays, previous versions of FoundationOne Liquid CDx. **Table 37** below includes a summary of the tissue types included in the study. Overall, 98.1% of samples yielded  $\geq 25$ ng DNA, which corresponds to a DNA input mass of 20ng for library construction (LC). A total of 89.1% of samples yielded  $\geq 36$ ng of DNA which corresponds to a DNA input mass of 30ng for LC. The proportion of samples with an LC yield greater than the minimum mass of 500ng was 99.9%, with one sided 95% confidence interval of [99.8%, 99.9%]. The proportion of samples with an HC yield greater than the minimum mass of 1000ng was 100%, with one sided 95% confidence interval of [99.99%, 100%]. The proportion of samples which met coverage requirements was 96.2%, with one sided 95% confidence interval of [95.9%, 96.3%]. The proportion of samples that generated a passing or qualified result after sequencing was 95.4%, with one sided 95% confidence interval of [95.1%, 95.6%].

**Table 37. F1L/FACT samples per tumor type and pass rates**

Tumor Type	Sample Size	DNA Extraction Pass Rate ( $\geq 25$ ng <sup>2</sup> )	DNA Extraction Pass Rate ( $\geq 36$ ng <sup>1</sup> )	LC Yield Pass Rate	HC Yield Pass Rate	Median Coverage Pass Rate	Overall Pass Rate ( $\geq 36$ ng <sup>1</sup> )	Overall Pass Rate ( $\geq 25$ ng <sup>2</sup> )
Rare Tumors	1164	97.0%	86.4%	99.9%	100.0%	93.8%	94.0%	97.0%
Biliary Cancer	171	99.4%	95.3%	100.0%	100.0%	98.8%	97.1%	99.4%
Bladder Cancer	166	97.6%	85.5%	100.0%	100.0%	93.2%	98.8%	97.6%
Breast Cancer	2775	97.6%	87.7%	99.9%	100.0%	96.4%	95.3%	97.6%
Cholangiocarcinoma	377	98.9%	96.0%	99.7%	100.0%	98.7%	96.8%	98.9%
Colorectal Cancer (CRC)	1640	98.5%	92.4%	99.9%	100.0%	97.5%	96.9%	98.5%

Endocrine-Neuro Cancer	75	100.0%	85.3%	100.0%	100.0%	100.0%	93.3%	100.0%
Endometrial Cancer	231	98.3%	88.3%	100.0%	100.0%	96.5%	95.6%	98.3%
Esophagus Cancer	291	99.7%	92.4%	100.0%	100.0%	97.6%	96.6%	99.7%
Glioma Cancer	59	94.9%	72.9%	100.0%	100.0%	100.0%	76.8%	94.9%
Head and Neck Cancer	154	96.1%	81.8%	100.0%	100.0%	89.2%	95.3%	96.1%
Kidney Cancer	203	99.0%	87.7%	100.0%	100.0%	95.0%	95.0%	99.0%
Liver Cancer	109	98.2%	95.4%	100.0%	100.0%	100.0%	95.3%	98.2%
Lung Non-Small Cell Lung Carcinoma (NSCLC)	5919	98.2%	88.8%	99.8%	100.0%	95.5%	95.4%	98.2%
Melanoma	257	96.5%	79.8%	100.0%	100.0%	92.7%	93.1%	96.5%
Ovary Cancer	496	97.8%	88.5%	100.0%	100.0%	95.9%	94.2%	97.8%
Pancreas Cancer	1359	98.8%	94.0%	99.9%	100.0%	97.8%	95.5%	98.8%
Peripheral Nervous System (PNS)	44	100.0%	90.9%	100.0%	100.0%	100.0%	93.2%	100.0%
Prostate Cancer	1778	97.3%	87.7%	99.9%	100.0%	96.9%	95.1%	97.3%
Small Cell Cancer	135	98.5%	93.3%	100.0%	100.0%	99.2%	99.2%	98.5%
Soft Tissue Sarcoma	130	97.7%	83.1%	100.0%	100.0%	95.3%	92.1%	97.7%
Stomach Cancer	267	98.9%	89.1%	100.0%	100.0%	98.1%	93.2%	98.9%
Thyroid Cancer	50	98.0%	86.0%	100.0%	100.0%	100.0%	81.6%	98.0%
Unspecified	856	98.5%	89.1%	100.0%	100.0%	95.5%	96.3%	98.5%
Unknown Primary Carcinoma (CUP)	1162	98.1%	89.7%	100.0%	100.0%	95.2%	95.7%	98.1%

<sup>1</sup> 36 ng of extracted cfDNA allows for sufficient cfDNA to process 30 ng of cfDNA

<sup>2</sup> 25 ng of extracted cfDNA allows for sufficient cfDNA to process 20 ng of cfDNA

**Table 38** summarizes the overall sample pass rate across tumor types as well as performance metrics from key QC points in the process. These results demonstrate comparable test performance across tumor types.

**Table 38: Summary of F1L/FACT sample data**

QC Metric	QC Pass Rate Across Tumor Types	Tumor Types with $\geq 90\%$ QC Pass Rate
Overall report Pass/Qualified rate	76.8%~99.2%	23/25 (92%)
Library Construction	99.7%~100%	25/25 (100%)
Hybridization Capture	100%	25/25 (100%)
Median exon coverage	89.2%~100%	24/25 (96%)

### 9.16 Concordance – FoundationOne Liquid Laboratory Developed Test to FoundationOne Liquid CDx

In order to support the use of historical data from the FoundationOne Liquid LDT to evaluate performance across cancer types, a study was performed to evaluate concordance between FoundationOne Liquid CDx and the FoundationOne Liquid LDT across the genomic regions targeted by both assays. This study evaluated the concordance of 927 unique samples processed on both the FoundationOne Liquid laboratory developed test (LDT) and FoundationOne Liquid CDx assays. A total of 3,366 alterations, consisting of only those in common between the assays were evaluated. The concordance analysis using FoundationOne Liquid LDT or FoundationOne Liquid CDx as the reference assay is summarized by variant category in **Table 39**.



**Table 39. Concordance between FoundationOne Liquid LDT (F1L LDT) and FoundationOne Liquid CDx (F1L CDx)**

Variant/ Mutation Type	F1L CDx+ F1L LDT+	F1L CDx- F1L LDT+	F1L CDx+ F1L LDT-	F1L CDx- F1L LDT -	PPA [95% CI]	NPA [95% CI]	OPA [95% CI]
<b>All Short Variants</b>	2871	123	32	1171180	95.9% [95.1%-96.6%]	>99.9% [>99.9%-100.0%]	>99.9% [>99.9%-100.0%]
<b>Base Substitutions</b>	2415	104	31	999032	95.9% [95.0%-96.6%]	>99.9% [>99.9%-100.0%]	>99.9% [>99.9%-100.0%]
<b>Indels</b>	456	19	1	172148	96.0% [93.8%-97.6%]	>99.9% [>99.9%-100.0%]	>99.9% [>99.9%-100.0%]
<b>Rearrangements</b>	147	20	24	59587	88.0% [82.1%-92.5%]	>99.9% [>99.9%-100.0%]	99.9% [99.9%-99.9%]
<b>Copy Number Amplifications</b>	173	32	0	59463	84.4% [78.7%-89.1%]	99.8% [>99.9%-100.0%]	99.8% [>99.9%-100.0%]
<b>Total</b>	3191	175	166	1290230	94.8% [94.0%-95.5%]	>99.9% [>99.9%-100.0%]	>99.9% [>99.9%-100.0%]

The overall PPA between FoundationOne Liquid LDT and FoundationOne Liquid CDx assays, with FoundationOne Liquid LDT as the reference assay, was 94.8% with a 95% two-sided CI of [94.0%-95.5%]. The respective short variant, rearrangement, and copy number amplification PPA values, with 95% two-sided CI, were: 95.9% [95.1%-96.6%], 88.0% [82.1%-92.5%], and 84.4% [78.7%-89.1%]. These results support the agreement between FoundationOne Liquid LDT and FoundationOne Liquid CDx and the applicability of the tumor comparability analysis performed using historical FoundationOne Liquid data.

### 9.17 Molecular Index Barcode Performance

To evaluate the molecular index barcode performance, a total of 7,641 sequenced samples from FoundationOne Liquid CDx validation studies were analyzed with the FoundationOne Liquid CDx assay.

The overall coefficient of variation (% CV) of sequencing coverage across all barcodes was 8.95% for the enhanced sensitivity regions and 7.64% for the standard sensitivity regions. This observed small % CV includes both sample variability and barcode variability as these two components were confounded and inseparable. Results demonstrated that all 480 barcodes analyzed are detectable with low differences in sample coverage variance between barcodes, indicating comparable performance of the barcodes.

### 9.18 Automation Line Equivalence

An intermediate precision study was performed to establish equivalence between the Hamilton instrumentation and the Biomek/Bravo instrumentation. The study consisted of eight contrived samples run in triplicate across four runs and both instrumentation platforms resulting in a total of 192 sample replicates included in the study overall. The analysis evaluated the negative call rate (NCR) and positive call rate (PCR) for 1,309 variants from eight contrived samples. The PCR and NCR were also evaluated by the seven variant categories.

The Mann-Whitney test was used for the comparison of PCR and NCR across liquid handling platforms for each sample, all samples in aggregate, and for each variant type. The NCR across platforms for each analysis set (per sample, all samples in aggregate, per variant type) were not statistically significant ( $p > 0.05$ ). by sample and by variant type. The PCR across platforms were not statistically significant ( $p > 0.05$ ) with the exception of contrived sample #3, the aggregate of all samples, and substitutions in a non-repetitive region or a repetitive region of  $\leq 7$  base pairs. The PCRs for the Hamilton liquid handling platform were slightly higher than the PCRs for the Biomek/Bravo platform (92.08% versus 90.15% for sample #3, 90.75% versus 89.67% for all samples, and 91.14 versus 90.10% for substitutions in a non-repetitive region or repetitive region of  $\leq 7$  base pairs). The statistical significance observed was due to large sample sizes allowing for the detection of slight differences that are likely

not meaningful in practice; therefore, the Hamilton and Biomek/Bravo liquid handling platforms are considered to be interchangeable in the FoundationOne Liquid CDx assay.

## 10 Clinical Validation Studies

### 10.1 Clinical Bridging Study: Detection of *ALK* Rearrangements to Determine Eligibility for Treatment with Alectinib

The clinical validity of using FoundationOne Liquid CDx as a companion diagnostic to identify patients with non-small cell lung cancer (NSCLC) harboring *ALK* rearrangements for treatment with alectinib was assessed through a clinical bridging study using screening (i.e., pre-alectinib treatment) plasma samples from Cohort A of the Blood First Assay Screening Trial (BFAST, BO29554).

The BFAST trial is a Phase II/III multicenter study, in which Cohort A evaluated the safety and efficacy of alectinib as a treatment for patients with advanced or metastatic NSCLC who tested positive for an *ALK* rearrangement as determined by a blood-based NGS assay (CTA).

The concordance between FoundationOne Liquid CDx and the CTA was evaluated as summarized in **Table 40**.

**Table 40: Concordance between FoundationOne Liquid CDx and the CTA for the detection of *ALK* rearrangements**

	CTA Pos	CTA Neg	Total
FoundationOne Liquid CDx Positive	63	0	63
FoundationOne Liquid CDx Negative	12	174	186
Missing	4	9	13
Total	79	183	262

The Positive Percent Agreement (PPA) and Negative Percent Agreement (NPA) between FoundationOne Liquid CDx and the CTA using the CTA as the reference for the primary analysis set and the corresponding 95% confidence intervals were:

- PPA [95% CI]: 84.0% [73.7%, 91.4%]
- NPA [95% CI] : 100% [ 97.9%, 100.0%]

After adjusting for a 5% prevalence of *ALK* rearrangements in the intended use population, the Positive Predictive Value (PPV), and Negative Predictive Value (NPV) calculated using the CTA as the reference and the corresponding 95% confidence intervals were:

- PPV [95% CI]: 100.0% [94.3%, 200.0%]
- NPV [95% CI]: 93.5% [89.0%, 96.6%]

The estimated Overall Response Rate (ORR) and the corresponding 95% confidence intervals was 88.9% [78.4%, 95.4%] for the FoundationOne Liquid CDx *ALK*-positive population which is comparable with the observed ORR and the corresponding 95% confidence intervals of 87.4% [78.5%, 93.5%] for the CTA *ALK*- positive population (BFAST Cohort A).

A sensitivity analysis was performed to estimate the clinical efficacy of treating patients with alectinib when considering missing FoundationOne Liquid CDx results. The estimated ORR and the corresponding 95% confidence intervals were 90.4% [90.1%, 90.6%] for the patient population that are both CTA *ALK*+ and FoundationOne Liquid CDx *ALK*+, demonstrating the robustness of the clinical efficacy analysis to missing FoundationOne Liquid CDx results.

## 10.2 FoundationOne Liquid CDx Concordance Study for *EGFR* Exon 19 deletion and *EGFR* Exon 21 L858R Alteration

Clinical validity of FoundationOne Liquid CDx assay was established as a companion diagnostic to identify patients with advanced NSCLC who may be eligible for treatment with TARCEVA® (erlotinib), IRESSA® (gefitinib), or TAGRISSO® (osimertinib). Two hundred and eighty retrospective samples from NSCLC patients were included in this study, which were tested for *EGFR* exon 19 deletion and exon 21 L858R alterations (*EGFR* alterations) by the FoundationOne Liquid CDx assay and the previously approved **cobas®** *EGFR* Mutation Test v2 (Roche Molecular Systems, referred to cobas assay). Both *EGFR* alteration-positive and *EGFR* alteration-negative samples (based on CTA results) were selected from the screen failed population of an unrelated clinical trial in NSCLC. To avoid selection bias, the samples were selected starting with a specific testing date until the predefined number of 150 *EGFR* alteration-positive and 100 *EGFR* alteration-negative samples were fulfilled. Samples were tested across two replicates by the cobas assay (denoted as CCD1 and CCD2) and one replicate by FoundationOne Liquid CDx. The tested samples, from NSCLC patients, were compared against the intended use (IU) population with respect to gender to ensure the screening population is representative of the IU population. The variant calls were evaluated based on the agreement between both the FoundationOne Liquid CDx and the cobas assay results and between the two cobas assay replicates. For any samples in which there was insufficient plasma to process both CCD1 and CCD2, processing was not performed. In total there were 177 samples with complete test results available for analysis. The agreement analysis results between FoundationOne Liquid CDx and the cobas assay for the detection of *EGFR* exon 19 deletions and L858R alterations are presented in **Table 41**.

**Table 41: Agreement analysis results for *EGFR* exon 19 deletion and L858R separately.**

Exon 19 deletion	PPAC1F	95.5%	NPAC1F	95.6%
	PPAC1C2	97.7%	NPAC1C2	98.9%
	PPAC2F	95.5%	NPAC2F	96.0%
	PPAC2C1	96.2%	NPAC2C1	99.4%
L858R	PPAC1F	100.0%	NPAC1F	95.6%
	PPAC1C2	92.9%	NPAC1C2	98.9%
	PPAC2F	100.0%	NPAC2F	94.7%
	PPAC2C1	96.0%	NPAC2C1	98.0%

The concordance of *EGFR* mutations as detected by FoundationOne Liquid CDx and the cobas assay were assessed and the data are summarized in **Table 42**.

**Table 42: Concordance among CCD1, CCD2 and FoundationOne Liquid CDx results with eligible samples (n=177)**

	CCD1+			CCD1-		
	CCD2+	CCD2-	Total	CCD2+	CCD2-	Total
FoundationOne Liquid CDx+	80	4	84	1	3	4
FoundationOne Liquid CDx-	2	0	2	0	87	87
<b>Total</b>	82	4	86	1	90	91

The agreement analysis results between FoundationOne Liquid CDx and the cobas assay are presented in **Table 43**.

**Table 43: Agreement analysis results**

	PPA	NPA
CCD2 CCD1*	95.3%	98.9%

CCD1 CCD2**	96.1%	98.7%
FoundationOne Liquid CDx CCD1*	97.7%	95.6%
FoundationOne Liquid CDx  CCD2**	97.7%	95.4%

\* CCD1: the 1st replicate of cobas assay as the reference

\*\* CCD2: the 2nd replicate of cobas assay as the reference

The estimates of  $\zeta$ PPA1,  $\zeta$ PPA2,  $\zeta$ NPA1 and  $\zeta$ NPA2 and the corresponding one-sided 95% upper bounds confidence limit computed using the bootstrap method are presented in **Table 44**.

**Table 44: Point estimate and one-Sided 95% upper confidence limit of  $\zeta$ PPA1,  $\zeta$ NPA1,  $\zeta$ PPA2, and  $\zeta$ NPA**

	Point Estimate	Mean one-sided 95% upper confidence limit
$\zeta$ PPA1	-2.3%	2.3%
$\zeta$ NPA1	3.3%	6.6%
$\zeta$ PPA2	-1.6%	4.7%
$\zeta$ NPA2	3.3%	6.6%

Based on these results, FoundationOne Liquid CDx has been demonstrated to be non-inferior to the cobas assay for the detection of *EGFR* exon 19 deletions and *EGFR* exon 21 L858R mutations. This study establishes the clinical validity of the FoundationOne Liquid CDx assay for identifying patients eligible for treatment with erlotinib, gefitinib, and osimertinib.

### 10.3 Clinical Bridging Study: Detection of *BRCA1/BRCA2/ATM* Alterations to Determine Eligibility for Treatment with Olaparib

The clinical validity of using FoundationOne Liquid CDx as a companion diagnostic to identify patients with metastatic castrate-resistant prostate cancer (mCRPC) harboring *BRCA1*, *BRCA2* or *ATM* alterations for treatment with olaparib was assessed through a clinical bridging study using screening (i.e., pre-olaparib treatment) plasma samples from Cohort A of the PROfound trial.

The PROfound trial is a Phase III, open label, randomized study to assess the efficacy and safety of olaparib (Lynparza™) versus enzalutamide or abiraterone acetate in men with metastatic castration-resistant prostate cancer who have failed prior treatment with a new hormonal agent and have homologous recombination repair gene mutations. Only Cohort A patients with either *BRCA1*, *BRCA2* or *ATM* mutations were tested with the FoundationOne Liquid CDx assay.

In total, 4425 patients were screened and 387 (9.6%) were randomized into the PROfound study by the CTA. Of these 387 patients, 245 patients were randomized in cohort A. 181 out of the 245 randomized patients in cohort A both consented to the use of their sample for ctDNA CDx development and had a plasma sample available for testing. In total, 181/245 (73.9%) of the Cohort A patients were tested using the FoundationOne Liquid CDx assay. Of these, 139 (76.8%) Cohort A patients had a successful FoundationOne Liquid CDx test result and 42 Cohort A patients had a failed FoundationOne Liquid CDx test result. This represents 56.7% (139/245) of total Cohort A patients with a FoundationOne Liquid CDx result. In addition, 250 non-HRRm patient samples were randomly selected for ctDNA testing from the screen-failed population to determine the NPA/NPV of the FoundationOne Liquid CDx assay. 194/250 (77.6%) screen failed non-HRRm patients were successfully tested using the FoundationOne Liquid CDx assay.

Of the 139 successfully tested Cohort A patients, 111 patients were reported as *BRCA1/BRCA2/ATM* mutation positive and 28 randomized patients were reported as biomarker negative by FoundationOne Liquid CDx.

Therefore, the FoundationOne Liquid CDx ctDNA biomarker positive subgroup comprises 111 patients with *BRCA1*, *BRCA2*, and/or *ATM* mutations.

Sample accountability for this clinical bridging study is summarized in **Table 45**.

**Table 45: Sample accountability for olaparib clinical bridging study**

Description	Number of patients
Patients randomized into PROfound	387
Patients with qualifying <i>BRCA1</i> , <i>BRCA2</i> , or <i>ATM</i> alterations (Cohort A)	245
Cohort A patients with samples tested by FoundationOne Liquid CDx	181
FoundationOne Liquid CDx results available	139
Cohort A patients, biomarker positive by FoundationOne Liquid CDx	111

**Table 46** shows the agreement analysis between CLIA CTA (tissue test) and the FoundationOne® Liquid CDx results for PROfound patients, including Invalid and Not Tested results

**Table 46: Summary of agreement analyses for FoundationOne® Liquid CDx compared against CTA tissue test, including Invalid and Not Tested results**

		CTA Results (n=495)	
		Biomarker <sup>a</sup> positive	Biomarker <sup>a</sup> negative
FMI F1 Liquid CDx assay	Biomarker <sup>a</sup> positive	111	16
	Biomarker <sup>a</sup> negative	28	178
	Biomarker <sup>a</sup> Invalid	42	56
	Not Tested	64	0
Agreement analyses (only Valid results included)	PPA (95% CI <sup>b</sup> )	79.9 (72.2, 86.2) [111/139]	
	NPA (95% CI <sup>b</sup> )	91.8 (87.0, 95.2) [178/194]	
	OPA (95% CI <sup>b</sup> )	86.8 (82.7, 90.2) [289/333]	
	PPV (95% CI <sup>b</sup> )	66.6 (56.0, 77.2)	
	NPV (95% CI <sup>b</sup> )	95.7 (94.3, 97.1)	

<sup>a</sup> Biomarker refers to patients with eligible *BRCA/ATM* mutations

<sup>b</sup> Confidence intervals calculated using Clopper-Pearson method

The Positive Percent Agreement (PPA) and Negative Percent Agreement (NPA) between FoundationOne Liquid CDx and the CTA using the CTA as the reference for the primary analysis set and the corresponding 95% confidence intervals were:

- PPA [95% CI]: 79.9% [72.2%, 86.2%]
- NPA [95% CI] : 91.8% [87.0%, 95.2%]

After adjusting for a 17.1% prevalence of BRCA1/2 and ATM alterations in the intended use population, the Positive Predictive Value (PPV), and Negative Predictive Value (NPV) calculated using the CTA as the reference and the corresponding 95% confidence intervals were:

- PPV [95% CI]: 66.6% [56.0%, 77.2%]
- NPV [95% CI]: 95.7% [94.3%, 97.1%]

The estimated radiological progression-free survival (rPFS) hazard ratio (HR) and the corresponding 95% confidence intervals were 0.331 [0.21, 0.53] for the FoundationOne Liquid CDx biomarker positive population, which were comparable with the observed rPFS HR and the corresponding 95% confidence intervals of 0.34 [0.25, 0.47] for the CTA biomarker positive population (PROfound Cohort A).

Sensitivity analysis to evaluate the robustness of the clinical efficacy estimate against the unknown FoundationOne Liquid CDx results was performed using the multiple imputation method in All Patients. After imputing the missing FoundationOne Liquid CDx results, the median rPFS HR and corresponding [95% CI] across the imputed datasets was 0.44 [0.32, 0.59], demonstrating robustness of the analysis to missing FoundationOne Liquid CDx results.


#### 10.4 Clinical Bridging Study: Detection of BRCA1 and BRCA2 Alterations to Determine Eligibility of mCRPC Patients for Treatment with Rucaparib

The clinical performance of FoundationOne Liquid CDx as a companion diagnostic to identify patients with metastatic castration-resistant prostate cancer (mCRPC) harboring breast cancer gene 1 or 2 (BRCA1 or BRCA2) alterations for treatment with rucaparib was demonstrated using pre-rucaparib treatment blood samples from clinical trial NCT0952534 (TRITON2). The clinical data supporting the use of rucaparib in the proposed indication was submitted as New Drug Application (NDA) 209115/S-004.

A bridging study was conducted to evaluate: 1) the concordance between BRCA1 and BRCA2 alteration status by the clinical trial assay (CTA) and FoundationOne Liquid CDx, and 2) the clinical efficacy of rucaparib treatment in patients that would be eligible for therapy based on BRCA1 and BRCA2 alteration status as determined by FoundationOne Liquid CDx.

A total of 209 patients (All Patients) from TRITON2 were included in NDA 209115/S-004. Genomic status was determined using the FoundationOne laboratory developed test [LDT] (F1 LDT), the FoundationOne Liquid LDT (F1L LDT), or a local test, as summarized in **Figure 1**.

**Figure 1: TRITON2 Patient Enrollment**



Pre-rucaparib treatment plasma samples were available for 92% (192/209) of the patients. FoundationOne Liquid CDx data were available for 93% (178/192) of the patients with samples tested; inadequate input material resulted in FoundationOne Liquid CDx test data being unavailable for 14 patients. In total, FoundationOne Liquid CDx data were available for 85% (178/209) of All Patients.

Of the 62 patients in the Primary Efficacy Population (those patients with measurable visceral and/or nodal disease at baseline), FoundationOne Liquid CDx test data were obtained for 84% (52/62) and used for concordance and efficacy analyses. The sample accountability for this clinical bridging study is summarized in **Table 47**.

**Table 45: Sample accountability for rucaparib prostate clinical bridging study**

Description	Number
All Patients in TRITON2	209
Total samples available for retesting by FoundationOne Liquid CDx	192
Patients with evaluable FoundationOne Liquid CDx data and cfDNA input $\geq$ 30ng (All Patients)	161
Patients with evaluable FoundationOne Liquid CDx test results and cfDNA input $\geq$ 20ng (All Patients)	178
Primary efficacy population in TRITON2	62
Patients with evaluable FoundationOne Liquid CDx test results and cfDNA input $\geq$ 30ng (Primary Efficacy Population)	48
Patients with evaluable FoundationOne Liquid CDx test results and cfDNA input $\geq$ 20ng (Primary Efficacy Population)	52

### Concordance between FoundationOne Liquid CDx and the CTAs

The concordance of BRCA status between FoundationOne Liquid CDx and CTA test results were evaluated in all patients as summarized in **Table 48** and **Table 49**.

**Table 46: Concordance between FoundationOne Liquid CDx BRCA Status and the CTA BRCA Status in All Patients with FoundationOne Liquid CDx cfDNA input  $\geq$ 30ng**

All Patients		CTA		
		BRCA Positive	BRCA Negative	Total
FoundationOne Liquid CDx	BRCA Positive	75	1	76
	BRCA Negative	16	69	85
	BRCA Unknown	2	1	3
	Total	93	71	164

The PPA, NPA between FoundationOne Liquid CDx and the CTA, based on a cfDNA input  $\geq$ 30ng, were determined using the CTA as the reference for all patients.

PPA (95% CI): 82.4% (73.0%, 89.6%)  
 NPA (95% CI): 98.6% (92.3%, 100.0%)

**Table 47: Concordance between FoundationOne Liquid CDx BRCA Status and the CTA BRCA Status in All Patients with FoundationOne Liquid CDx cfDNA input  $\geq$ 20ng**

All Patients		CTA		
		BRCA Positive	BRCA Negative	Total
FoundationOne Liquid CDx	BRCA Positive	82	1	83
	BRCA Negative	18	77	95
	BRCA Unknown	3	2	5
	Total	103	80	183

The PPA, NPA between FoundationOne Liquid CDx and the CTA, based on a cfDNA input  $\geq$ 20ng, were determined using the CTA as the reference for all patients.

- PPA [95% CI]: 82.0% [73.1%, 89.0%]
- NPA [95% CI]: 98.7% [93.1%, 100%]

### Efficacy Based on FoundationOne Liquid CDx Results

*BRCA1* and *BRCA2* alteration status were verified retrospectively by FoundationOne Liquid CDx in 66% (41/62) of the patients in the Primary Efficacy Population. The ORR [95% CI] in the Primary Efficacy Population was 46.3% [30.7%-62.6%] in *BRCA* positive patients determined by FoundationOne Liquid CDx, which is comparable to the ORR of 43.5% [31.0%-56.7%] in patients identified by CTA (Table 50).

**Table 48: ORR in the primary efficacy population by CTA and FoundationOne Liquid CDx test results**

Primary Efficacy Population	FoundationOne Liquid CDx		CTA
	BRCA Positive N=38 ( $\geq$ 30 ng cfDNA input)	BRCA Positive N = 41 ( $\geq$ 20 ng cfDNA input)	BRCA Positive N = 62
Confirmed ORR (CR + PR), n (%)	18 (47.4)	19 (46.3)	27 (43.5)
95% CI(%)	31.0 – 64.2	30.7 - 62.6	31.0 – 56.7

Abbreviations: *BRCA* = breast cancer gene, includes *BRCA1* and *BRCA2*; CI = confidence interval; CTA = clinical trial assay; ORR = objective response rate; CR = complete response; PR = partial response.

Sensitivity analysis to evaluate the robustness of the clinical efficacy estimate against the unknown FoundationOne Liquid CDx results was performed using the multiple imputation method and demonstrated that the drug efficacy in the FoundationOne Liquid CDx positive population was robust to missing FoundationOne Liquid CDx results.

### 10.5 Clinical Validation Study: Detection of *BRCA1* and *BRCA2* Alterations to Determine Eligibility of Ovarian Cancer Patients for Treatment with Rucaparib

The clinical performance of FoundationOne Liquid CDx as a companion diagnostic to identify patients with ovarian cancer harboring *BRCA1* or *BRCA2* alterations for treatment with rucaparib was demonstrated using pre- rucaparib treatment blood samples from the ARIEL2 study.

The bridging study was conducted to evaluate: 1) the concordance between *BRCA1* and *BRCA2* alteration status by the CTA and FoundationOne Liquid CDx, and 2) the clinical efficacy of rucaparib treatment in patients that would be eligible for therapy based on *BRCA1* and *BRCA2* alteration status as determined by FoundationOne Liquid CDx.

The ARIEL2 study is complete and enrolled 491 patients (All Patients). Pre-rucaparib treatment plasma samples were available for 55% (271/491) of patients dosed in ARIEL2. FoundationOne Liquid CDx data were available for



80% (217/271) of the patients with samples tested; 49 failures were due to insufficient remaining plasma volume or insufficient DNA extraction yield. In total, FoundationOne Liquid CDx results were available for 44% (217/491) of All Patients.

Of the 64 patients in the primary efficacy population, FoundationOne Liquid CDx results were available for 42% (27/64) and used for concordance and efficacy analyses. The sample accountability for this clinical validation study is summarized in **Table 51**.

**Table 49: Sample accountability for rucaparib ovarian clinical bridging study**

Description	Number
All Patients	491
Total samples available	271
Patients with FoundationOne Liquid CDx data (All Patients)	217
Patients with FoundationOne Liquid CDx data (Primary Efficacy Population)	27

The concordance between FoundationOne Liquid CDx and CTA test results was evaluated in All Patients and is summarized in **Table 52**. The Primary Efficacy Population is summarized in **Table 53**.

**Table 50: Concordance between FoundationOne Liquid CDx and the CTA for the detection of BRCA1 or BRCA2 alterations in All Patients**

	CTA Positive	CTA Negative	Total
FoundationOne Liquid CDx Positive	60	4	64
FoundationOne Liquid CDx Negative	4	149	153
Missing	60	214	274
Total	124	367	491

The PPA and NPA between FoundationOne Liquid CDx and the CTA were determined using the CTA as the reference for All Patients:

- PPA [95% CI]: 93.8% [84.8%, 98.3%]
- NPA [95% CI]: 97.4% [93.4%, 99.3%]

**Table 51: Concordance between FoundationOne Liquid CDx and the CTA for the detection of BRCA1 or BRCA2 alterations in the primary efficacy population**

	CTA Positive	CTA Negative	Total
FoundationOne Liquid CDx Positive	26	0	26
FoundationOne Liquid CDx Negative	0	1	1
Missing	35	2	37
Total	61	3	64

The PPA and NPA between FoundationOne Liquid CDx and the CTA were determined using the CTA as the reference for the Primary Efficacy Population:

- PPA [95% CI]: 100% [86.8%, 100.0%]
- NPA [95% CI]: 100% [ 2.5%, 100.0%]

*BRCA1* and *BRCA2* alteration status was verified retrospectively by FoundationOne Liquid CDx in 41% (26/64) of the patients in the Primary Efficacy Population.

The ORR [95% CI] in the primary efficacy population was 53.8% [33.4%-73.4%] in *BRCA* Positive patients as determined by FoundationOne Liquid CDx, which is comparable to the ORR of 54.1% [40.8%-66.9%] in patients identified by the CTA (**Table 54**).

The median DOR [95% CI] was 225 days [115, 403] in FoundationOne Liquid CDx *BRCA* Positive patients from the Primary Efficacy Population. This is similar to the median DOR of 288 days [170, 403] for the Primary Efficacy Population in *BRCA* Positive patients by the CTA.

**Table 52: ORR and duration of response in the primary efficacy population by CTA and FoundationOne Liquid CDx test results**

	FoundationOne Liquid CDx <i>BRCA</i> Positive n = 26	CTA <i>BRCA</i> Positive n = 61
<b>Confirmed ORR (CR + PR), % (n)</b>	53.8% (14)	54.1% (33)
95% CI	33.4%, 73.4%	40.8%, 66.9%
<b>Duration of Response (days)</b>		
Median	225	288
95% CI	115 – 403	170 – 403

Abbreviations: *BRCA* = breast cancer gene, includes *BRCA1* and *BRCA2*; CI = confidence interval; CTA = clinical trial assay; ORR = objective response rate; CR = complete response; PR = partial response.

The ORR [95% CI] in All Patients was evaluated for *BRCA* Positive and *BRCA* Negative patients. The ORR in *BRCA* Positive patients identified from FoundationOne Liquid CDx was 40.6% [28.5%-53.6%] compared to the ORR of 46.8% [37.8%-55.9%] in *BRCA* Positive patients based on the CTA. The ORR in *BRCA* Negative patients by FoundationOne Liquid CDx and the CTA was 5.9% [2.7%-10.9%] and 13.1% [9.8%-17.0%], respectively.

Sensitivity analysis to evaluate the robustness of the clinical efficacy estimate against the unknown FoundationOne Liquid CDx results was performed using the multiple imputation method in All Patients. After imputing the missing FoundationOne Liquid CDx results, the weighted ORR [95% CI] across the imputed datasets was 45.2% [36.3%-54.1%].

## 10.6 Clinical Bridging Study: Detection of *PIK3CA* Alterations to Determine Eligibility for Treatment with Alpelisib

Clinical validity of using FoundationOne Liquid CDx to identify breast cancer patients harboring *PIK3CA* alterations eligible for treatment with alpelisib was assessed through retrospective testing of plasma samples collected prior to study treatment from advanced or metastatic breast cancer patients enrolled in clinical trial CBYL719C2301 (SOLAR-1). A total of 395 patients were enrolled based on CTA1 results and 177 patients were enrolled based on CTA2 results. All 395 patients enrolled based on CTA1 results were retrospectively tested by CTA2. This clinical bridging study was performed based on CTA2 results.

Samples with  $\geq 30$  ng from 375 patients were tested by FoundationOne Liquid CDx. Excluding those with invalid results for either CTA2 or CDx (4 and 12, respectively), the primary efficacy analyses were conducted using data

from the 359 subjects who were CTA2-evaluable and CDx-evaluable. A concordance analysis was conducted with the CTA2-evaluable and FoundationOne Liquid CDx-evaluable samples as summarized in **Table 55**.

**Table 53: Concordance between FoundationOne Liquid CDx and CTA2**

CDx	CTA2			Total
	+	-	Invalid	
+	165	0	1	166
-	65	129	3	197
Invalid	7	5	0	12
Total	237	134	4	375

Samples not tested are excluded from the analysis.

Samples tested with cfDNA input < 30 ng are excluded from the analysis.

The point estimates of PPA and NPA between FoundationOne Liquid CDx and the CTA2 assay and the corresponding 95% confidence intervals were:

- PPA [95% CI]: 71.7% [65.4%, 77.5%]
- NPA [95% CI]: 100% [97.2%, 100%]

The primary efficacy analysis in the *PIK3CA* alteration positive population identified by FoundationOne Liquid CDx was based on PFS by local investigator assessment per RECIST 1.1 criteria. Clinical efficacy of alpelisib in combination with fulvestrant for the FoundationOne Liquid CDx-positive population with cfDNA input  $\geq 30$  ng (N=165) was demonstrated with an estimated 54% risk reduction in disease progression or death in the alpelisib plus fulvestrant arm compared to the placebo plus fulvestrant arm (HR = 0.46, 95% CI: 0.30, 0.70).

As summarized in **Table 57**, the PFS hazard ratio for the 165 tissue CTA2-positive, FoundationOne Liquid CDx-positive patients was 0.46 (95% CI: 0.30, 0.70). Median PFS was 11.0 months for the alpelisib plus fulvestrant arm versus 3.6 months for the placebo plus fulvestrant arm.

**Table 57: Progression-free survival in the CTA2-positive, FoundationOne Liquid CDx-positive patients (primary analysis set)**

Progression free survival (months)	Alpelisib 300mg qd + Fulvestrant N=84	Placebo qd + Fulvestrant N=81	HR (95% CI) Alpelisib 300mg qd + Fulv /Placebo qd + Fulv <sup>1</sup>
No of events (%)	54 (64.3)	67 (82.7)	0.46 (0.30, 0.70)
PD (%)	52 (61.9)	61 (75.3)	
Death (%)	2 (2.4)	6 (7.4)	
No of censored (%)	30 (35.7)	14 (17.3)	
Median (95% CI) <sup>2</sup>	11.0 (7.3, 15.9)	3.6 (2.4, 5.8)	

<sup>1</sup> Hazard ratio (HR) estimated using Cox regression model stratified by the two stratification factors: presence of lung and/or liver metastases, previous treatment with any CDK4/6 inhibitor, and adjusted for clinically relevant covariates, as well as the imbalanced covariates.

<sup>2</sup> The 95% CI calculated from PROC LIFETEST output using the method of Brookmeyer and Crowley (1982).

CDx results from samples tested with cfDNA input < 30 ng are treated as missing.

PD = progressive disease

Sensitivity analysis to evaluate the robustness of the clinical efficacy estimate against the missing FoundationOne Liquid CDx results was performed using the multivariate imputation by chained equations (MICE) method. After imputing the missing FoundationOne Liquid CDx results, the hazard ratio was estimated to be 0.63 (95% CI: 0.45, 0.87), demonstrating robustness of the clinical efficacy analysis to missing FoundationOne Liquid CDx results.

## 11 CDx Classification Criteria

### 11.1 CDx classification criteria for *ALK* rearrangements, qualifying NSCLC patients for therapy with ALECENSA® (alectinib):

- The *ALK* rearrangement must have pathogenic driver status (FMI driver status of “known” or “likely”)
- AND the disease type must be NSCLC
- AND one of the following two conditions must hold:
  1. The partner gene is *EML4*, or
  2. The *ALK* breakpoint occurs within *ALK* intron 19

### 11.2 CDx classification criteria for *EGFR* alterations, qualifying NSCLC patients for therapy with IRESSA® (gefitinib), TAGRISSO® (osimertinib), TARCEVA® (erlotinib):

- Base substitutions resulting in *EGFR* L858R
- In-frame deletions occurring within *EGFR* exon 19

### 11.3 CDx classification criteria for *BRCA1*, *BRCA2*, and *ATM* alterations, qualifying prostate cancer patients for therapy with LYNPARZA® (olaparib):

Table 57, Table 58, and Table 59 describe the criteria for classifying *BRCA1*, *BRCA2*, or *ATM* alterations known to be deleterious to protein function

**Table 57: Classification Criteria for *BRCA1*, *BRCA2*, and *ATM***

Deleterious Variant Criteria	Sequence Classification	CDx Classifier Methodology
A gene alteration that includes any of the sequence classifications	Protein truncating mutations	Sequence analysis identifies premature stop codons or frameshift indels anywhere in the gene coding region, except: 3' of and including <i>BRCA2</i> K3326*
	Splice site mutations	Sequence analysis identifies variant splice sequences at intron/exon junctions: within $\pm$ 2bp of exon starts/ends, or callable splice variants in Table 59
	Homozygous deletions	Sequence analysis identifies deletions in both gene alleles of $\geq$ 1 exon in size.  Only reported for <i>BRCA1</i> and <i>BRCA2</i> . Not reported for <i>ATM</i> .
	Large protein truncating rearrangements	Sequence analysis identifies protein truncating rearrangements
	Deleterious missense mutations	Curated list (Table 58)

**Table 58: Deleterious Missense Alterations**

<i>BRCA1</i> Protein Effect (PE)	<i>BRCA2</i> Protein Effect (PE)	<i>ATM</i> Protein Effect (PE)
M1V	M1R	M1T
M1I	M1I	R2032K
C61G	V159M	R2227C

<b>BRCA1 Protein Effect (PE)</b>	<b>BRCA2 Protein Effect (PE)</b>	<b>ATM Protein Effect (PE)</b>
C64Y	V211L	R2547 S2549del
R71G	V211I	G2765S
R71K	R2336P	R2832C
R1495M	R2336H	S2855 V2856delinsRI
E1559K		R3008C
D1692N		R3008H
D1692H		
R1699W		
A1708E		
G1788V		

**Table 59: Intronic Variants**

Gene	Chromosome	Position	Ref	Alt	dbSNP
<i>ATM</i>	chr11	108128198	T	G	rs730881346
<i>ATM</i>	chr11	108214102	AGTGA	A	rs730881295

#### 11.4 CDx classification criteria for *BRCA1* and *BRCA2* alterations, qualifying prostate cancer or ovarian cancer patients for therapy with RUBRACA® (rucaparib):

**Table 60** and **Table 61** describe the criteria for classifying *BRCA1* or *BRCA2* alterations known to be deleterious to BRCA protein function rendering the sample *BRCA+*.

**Table 60: Classification Criteria for Deleterious Tumor *BRCA* Variants**

Qualification Criteria	Sequence Classification	Methodology
A <i>BRCA1/2</i> alteration that includes any of the sequence classifications	Protein truncating mutations	Sequence analysis identifies premature stop codons anywhere in the gene coding region, except: 3' of and including <i>BRCA2</i> K3326*
	Splice site mutations	Sequence analysis identifies variant splice sequences at intron/exon junctions +/- 2bp of exon starts/ends
	Homozygous deletions	Sequence analysis identifies deletions in both gene alleles of ≥ 1 exon in size
	Large protein truncating rearrangements	Sequence analysis identifies protein truncating rearrangements
	Deleterious missense mutations	Curated list ( <b>Table 61</b> )

**Table 61: Deleterious *BRCA* Missense Alterations**

<b>BRCA1 Alterations (Protein Change)</b>				<b>BRCA2 Alterations (Protein Change)</b>	
M1V	C61G	D1692H	G1788V	M1V	R2659T
M1T	C61Y	D1692Y	P1812A	M1T	R2659K
M1R	C64R	R1699W	A1823T	M1R	E2663V

<b>BRCA1 Alterations (Protein Change)</b>				<b>BRCA2 Alterations (Protein Change)</b>	
M1I	C64G	R1699Q	V1833M	M1I	S2670L
M18T	C64Y	G1706R	W1837R	D23N	I2675V
L22S	C64W	G1706E	V1838E	D23Y	T2722K
I26N	R71G	A1708E		S142N	T2722R
T37K	R71K	S1715R		S142I	D2723H
C39R	R71T	S1722F		V159M	D2723G
C39G	R71M	V1736A		V211I	G2724W
C39Y	S770L	G1738R		V211L	G2748D
C39W	R1495T	G1738E		Y600C	A2911E
H41R	R1495M	K1759N		K1530N	E3002K
C44S	R1495K	L1764P		R2336P	R3052W
C44Y	E1559K	I1766N		R2336L	D3095G
C44F	E1559Q	I1766S		R2336H	D3095E
C47S	T1685A	G1770V		T2412I	N3124I
C47Y	T1685I	M1775K		R2602T	N3187K
C47F	D1692N	M1775R		W2626C	
C61S	M1689R	C1787S		I2627F	

**11.5 CDx classification criteria for PIK3CA alterations, qualifying breast cancer patients for therapy with PIQRAY® (alpelisib):**

Presence of PIK3CA mutation(s): H1047R; E545K; E542K; C420R; E545A; E545D [1635G>T only]; E545G; Q546E; Q546R; H1047L; or H1047Y